



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114012

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	04/16/2021	13:02
2	CAL2	1	Water	NA	NA	21.3	7.00	04/16/2021	13:03
3	CAL3	1	Water	NA	NA	21.3	10.02	04/16/2021	13:04
4	ICV	1	Water	NA	NA	21.3	6.99	04/16/2021	13:05
5	CCV1	1	Water	NA	NA	21.3	2.01	04/16/2021	13:06
6	M2043-02	1	Water	NA	NA	20.3	7.02	04/16/2021	13:25
7	M2053-03	1	Water	NA	NA	20.6	6.92	04/16/2021	13:26
8	M2053-05	1	Water	NA	NA	20.6	6.90	04/16/2021	13:27
9	M2053-06	1	Water	NA	NA	20.6	7.55	04/16/2021	13:29
10	M2053-07	1	Water	NA	NA	20.2	7.23	04/16/2021	13:30
11	M2066-01	1	Water	NA	NA	20.4	4.77	04/16/2021	13:32
12	M2069-01	1	Water	NA	NA	20.5	9.31	04/16/2021	13:35
13	M2069-01DUP	1	Water	NA	NA	20.6	9.32	04/16/2021	13:36
14	CCV2	1	Water	NA	NA	20.3	12.02	04/16/2021	13:38

WORKLIST(Hardcopy Internal Chain)

203 114012

WorkList Name : PH M2066 WorkList ID : 148093 Department : Wet-Chemistry Date : 04-16-2021 12:53:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2021	Water	M2043-02	pH	Cool 4 deg C	URSC01	O51	FS-WC-WATER	04/14/2021	9040C
04/15/2021	Water	M2053-03	pH	Cool 4 deg C	PSEG03	O61	COMPOSITE-1	04/15/2021	9040C
04/15/2021	Water	M2053-05	pH	Cool 4 deg C	PSEG03	O62	GRAB-1	04/15/2021	9040C
04/15/2021	Water	M2053-06	pH	Cool 4 deg C	PSEG03	O62	GRAB-2	04/15/2021	9040C
04/15/2021	Water	M2053-07	pH	Cool 4 deg C	PSEG03	O62	GRAB-3	04/15/2021	9040C
04/15/2021	Water	M2066-01	pH	Cool 4 deg C	STER16	P11	EXCELL	04/15/2021	9040C
04/16/2021	Water	M2069-01	pH	Cool 4 deg C	TULL02	P12	MH20A-BNM-1-D-WATEF	04/16/2021	9040C

Date/Time 04-16-21 13:00
 Raw Sample Received by: JR (wet-deg)
 Raw Sample Relinquished by: CP (SM)

Date/Time 04-16-21 14:35
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: JR (wet-deg)