

LB114013


7471B

INSTRUMENT ID: CY1

Linear

μ Abs.:

53662



A= 0.0000e+000

B= 1.8612e-004

C= 4.6437e-003

Rho= 0.9999737

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	122	0.000	122				
0.2	0.200	0.206	0.006	1082	0.0 %	1082				
2.5	2.500	2.451	-0.049	13145	0.0 %	13145				
5.0	5.000	4.992	-0.008	26796	0.0 %	26796				
7.5	7.500	7.532	0.032	40442	0.0 %	40442				
10.0	10.000	9.992	-0.008	53662	0.0 %	53662				

0/011
1
3
2
000

LB114013 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Apr 2021 14:49:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	122	-	0.00007471B	PPB		16 Apr 2021 14:52:38	S	Std
0.2 - 1	50.2	1082	-	0.20007471B	PPB		16 Apr 2021 14:54:55	S	Std
2.5 - 1	52.5	13145	-	2.50007471B	PPB		16 Apr 2021 14:57:12	S	Std
5.0 - 1	55	26796	-	5.00007471B	PPB		16 Apr 2021 14:59:29	S	Std
7.5 - 1	57.5	40442	-	7.50007471B	PPB		16 Apr 2021 15:01:46	S	Std
10.0 - 1	51.6	53662	-	10.00007471B	PPB		16 Apr 2021 15:04:03	S	Std
ICV91 - 1	ICV91	21060	3.9243	-7471B	PPB		16 Apr 2021 15:06:47	U	SMPL
ICB91 - 1	ICB91	-66	-0.0076	-7471B	PPB		16 Apr 2021 15:09:02	U	SMPL
CCV28 - 1	CCV28	26714	4.9766	-7471B	PPB		16 Apr 2021 15:11:17	U	SMPL
CCB28 - 1	CCB28	-47	-0.0041	-7471B	PPB		16 Apr 2021 15:13:32	U	SMPL
CRA - 1	CRA	1040	0.1982	-7471B	PPB		16 Apr 2021 15:15:48	U	SMPL
HighStd - 1	HighStd	53473	9.9569	-7471B	PPB		16 Apr 2021 15:18:04	U	SMPL
ChkStd - 1	ChkStd	37742	7.0291	-7471B	PPB		16 Apr 2021 15:20:19	U	SMPL
PB135783BL - 1	PBS	-154	-0.0240	-7471B	PPB		16 Apr 2021 15:22:36	U	SMPL
M2050-01 - 1	TP-7	153788	28.6272	-7471B	PPB		16 Apr 2021 15:27:09	U	SMPL
M2050-03 - 1	TP-8	137736	25.6397	-7471B	PPB		16 Apr 2021 15:29:26	U	SMPL
M2050-05 - 1	TP-9	173180	32.2364	-7471B	PPB		16 Apr 2021 15:31:48	U	SMPL
M2050-07 - 1	TP-5	127729	23.7772	-7471B	PPB		16 Apr 2021 15:34:07	U	SMPL
M2050-09 - 1	TP-4	114701	21.3524	-7471B	PPB		16 Apr 2021 15:36:26	U	SMPL
CCV29 - 1	CCV29	26526	4.9416	-7471B	PPB		16 Apr 2021 15:38:45	U	SMPL
CCB29 - 1	CCB29	-113	-0.0164	-7471B	PPB		16 Apr 2021 15:41:04	U	SMPL
M2050-11 - 1	TP-3	169621	31.5740	-7471B	PPB		16 Apr 2021 15:43:20	U	SMPL
M2050-13 - 1	TP-2	78138	14.5475	-7471B	PPB		16 Apr 2021 15:45:35	U	SMPL
M2050-15 - 1	TP-1	243083	45.2465	-7471B	PPB		16 Apr 2021 15:47:53	U	SMPL
M2050-17 - 1	TP-6	51035	9.5031	-7471B	PPB		16 Apr 2021 15:50:11	U	SMPL
M2051-01 - 1	0124027	2300	0.4327	-7471B	PPB		16 Apr 2021 15:52:30	U	SMPL
M2051-01DUP - 1	0124027DUP	2399	0.4511	-7471B	PPB		16 Apr 2021 15:54:48	U	SMPL
M2051-01MS - 1	0124027MS	24142	4.4979	-7471B	PPB		16 Apr 2021 15:59:35	U	SMPL
M2051-01MSD - 1	0124027MSD	23873	4.4478	-7471B	PPB		16 Apr 2021 16:01:51	U	SMPL
M2051-03 - 1	0124035	7485	1.3977	-7471B	PPB		16 Apr 2021 16:04:08	U	SMPL
M2051-05 - 1	0124055	31090	5.7910	-7471B	PPB		16 Apr 2021 16:06:23	U	SMPL
CCV30 - 1	CCV30	27377	5.1000	-7471B	PPB		16 Apr 2021 16:08:38	U	SMPL
CCB30 - 1	CCB30	-24	0.0002	-7471B	PPB		16 Apr 2021 16:10:54	U	SMPL
M2051-07 - 1	0124019	19780	3.6860	-7471B	PPB		16 Apr 2021 16:13:10	U	SMPL
M2052-01 - 1	OK-02-041521	2515	0.4727	-7471B	PPB		16 Apr 2021 16:15:25	U	SMPL
M2053-01 - 1	PADS-2	1524	0.2883	-7471B	PPB		16 Apr 2021 16:17:40	U	SMPL
PB135783BS - 1	LCSS	22940	4.2742	-7471B	PPB		16 Apr 2021 16:19:55	U	SMPL
M2051-01LX5 - 1		435	0.0856	-7471B	PPB		16 Apr 2021 16:22:11	U	SMPL
M2051-01A - 1		25276	4.7089	-7471B	PPB		16 Apr 2021 16:24:27	U	SMPL
CCV31 - 1	CCV31	27746	5.1686	-7471B	PPB		16 Apr 2021 16:26:43	U	SMPL
CCB31 - 1	CCB31	-30	-0.0009	-7471B	PPB		16 Apr 2021 16:29:00	U	SMPL
M2050-01DLX5 - 1		32724	6.0951	-7471B	PPB		16 Apr 2021 16:31:17	U	SMPL
M2050-03DLX5 - 1		28383	5.2872	-7471B	PPB		16 Apr 2021 16:33:32	U	SMPL
M2050-05DLX5 - 1		36238	6.7492	-7471B	PPB		16 Apr 2021 16:35:49	U	SMPL
M2050-07DLX5 - 1		26059	4.8547	-7471B	PPB		16 Apr 2021 16:38:04	U	SMPL
M2050-09DLX5 - 1		23282	4.3378	-7471B	PPB		16 Apr 2021 16:40:21	U	SMPL
M2050-11DLX5 - 1		35996	6.7041	-7471B	PPB		16 Apr 2021 16:42:37	U	SMPL
M2050-13DLX5 - 1		15853	2.9552	-7471B	PPB		16 Apr 2021 16:44:52	U	SMPL
M2050-15DLX10 - 1		26348	4.9085	-7471B	PPB		16 Apr 2021 16:47:09	U	SMPL
CCV32 - 1	CCV32	27610	5.1433	-7471B	PPB		16 Apr 2021 16:49:25	U	SMPL
CCB32 - 1	CCB32	-39	-0.0026	-7471B	PPB		16 Apr 2021 16:51:41	U	SMPL