


LB114038

7470A

INSTRUMENT ID: CV2

 μ Abs.:

40864



Linear

A= 0.0000e+000

B= 2.4230e-004

C= -6.2566e-002

Rho= 0.9994402

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o))
0.0	0.000	0.001	0.001	264	0.000	264					7
0.2	0.200	0.212	0.012	1135	0.0 %	1135					6
2.5	2.500	2.432	-0.068	10297	0.0 %	10297					-3
5.0	5.000	4.970	-0.030	20770	0.0 %	20770					-1
7.5	7.500	7.745	0.245	32221	0.0 %	32221					3
10.0	10.000	9.839	-0.161	40864	0.0 %	40864					-2

LB114038 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2021 11:21:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	264	-	0.0000	7470A	PPB	19 Apr 2021 11:28:16	S	Std
0.2 - 1	50.2	1135	-	0.2000	7470A	PPB	19 Apr 2021 11:30:32	S	Std
2.5 - 1	52.5	10297	-	2.5000	7470A	PPB	19 Apr 2021 11:32:50	S	Std
5.0 - 1	55	20770	-	5.0000	7470A	PPB	19 Apr 2021 11:35:06	S	Std
7.5 - 1	57.5	32221	-	7.5000	7470A	PPB	19 Apr 2021 11:37:22	S	Std
10.0 - 1	510	40864	-	10.0000	7470A	PPB	19 Apr 2021 11:42:23	S	Std
ICV92 - 1	ICV92	15363	3.6599	-7470A	PPB	PPB	19 Apr 2021 11:45:06	U	SMPL
ICB92 - 1	ICB92	2	-0.0621	-7470A	PPB	PPB	19 Apr 2021 11:47:21	U	SMPL
CCV33 - 1	CCV33	19705	4.7120	-7470A	PPB	PPB	19 Apr 2021 11:49:36	U	SMPL
CCB33 - 1	CCB33	41	-0.0526	-7470A	PPB	PPB	19 Apr 2021 11:51:51	U	SMPL
CRA - 1	CRA	1027	0.1863	-7470A	PPB	PPB	19 Apr 2021 11:54:06	U	SMPL
HighStd - 1	HighStd	40576	9.7691	-7470A	PPB	PPB	19 Apr 2021 11:56:21	U	SMPL
ChkStd - 1	ChkStd	30160	7.2453	-7470A	PPB	PPB	19 Apr 2021 12:05:04	U	SMPL
PB135790BL - 1	PBW	1	-0.0623	-7470A	PPB	PPB	19 Apr 2021 12:07:19	U	SMPL
PB135790BS - 1	LCSW	15873	3.7835	-7470A	PPB	PPB	19 Apr 2021 12:12:06	U	SMPL
M2048-02 - 1	SB-01-COMP	65	-0.0468	-7470A	PPB	PPB	19 Apr 2021 12:14:21	U	SMPL
M2048-02DUP - 1	SB-01-COMPDUP	238	-0.0049	-7470A	PPB	PPB	19 Apr 2021 12:16:38	U	SMPL
M2048-02MS - 1	SB-01-COMPMS	15861	3.7806	-7470A	PPB	PPB	19 Apr 2021 12:21:06	U	SMPL
M2048-02MSD - 1	SB-01-COMPMSD	15869	3.7826	-7470A	PPB	PPB	19 Apr 2021 12:23:22	U	SMPL
M2048-05 - 1	SB-02-COMP	-12	-0.0655	-7470A	PPB	PPB	19 Apr 2021 12:25:37	U	SMPL
CCV34 - 1	CCV34	21151	5.0624	-7470A	PPB	PPB	19 Apr 2021 12:27:53	U	SMPL
CCB34 - 1	CCB34	51	-0.0502	-7470A	PPB	PPB	19 Apr 2021 12:30:07	U	SMPL
M2048-08 - 1	SB-03-COMP	122	-0.0330	-7470A	PPB	PPB	19 Apr 2021 12:32:22	U	SMPL
M2048-11 - 1	SB-04-COMP	190	-0.0165	-7470A	PPB	PPB	19 Apr 2021 12:34:39	U	SMPL
M2048-14 - 1	SB-05-COMP	214	-0.0107	-7470A	PPB	PPB	19 Apr 2021 12:36:54	U	SMPL
M2048-17 - 1	SB-06-COMP	225	-0.0080	-7470A	PPB	PPB	19 Apr 2021 12:39:10	U	SMPL
M2048-20 - 1	SB-07-COMP	158	-0.0243	-7470A	PPB	PPB	19 Apr 2021 12:41:25	U	SMPL
M2050-02 - 1	TP-7	190	-0.0165	-7470A	PPB	PPB	19 Apr 2021 12:43:41	U	SMPL
M2050-04 - 1	TP-8	235	-0.0056	-7470A	PPB	PPB	19 Apr 2021 12:45:57	U	SMPL
M2050-06 - 1	TP-9	193	-0.0158	-7470A	PPB	PPB	19 Apr 2021 12:48:14	U	SMPL
M2050-08 - 1	TP-5	271	0.0031	-7470A	PPB	PPB	19 Apr 2021 12:50:31	U	SMPL
M2050-10 - 1	TP-4	206	-0.0127	-7470A	PPB	PPB	19 Apr 2021 12:52:47	U	SMPL
CCV35 - 1	CCV35	20681	4.9485	-7470A	PPB	PPB	19 Apr 2021 12:55:02	U	SMPL
CCB35 - 1	CCB35	101	-0.0381	-7470A	PPB	PPB	19 Apr 2021 12:57:17	U	SMPL
M2050-12 - 1	TP-3	426	0.0407	-7470A	PPB	PPB	19 Apr 2021 12:59:33	U	SMPL
M2050-14 - 1	TP-2	296	0.0092	-7470A	PPB	PPB	19 Apr 2021 13:01:48	U	SMPL
M2050-16 - 1	TP-1	424	0.0402	-7470A	PPB	PPB	19 Apr 2021 13:04:03	U	SMPL
M2050-18 - 1	TP-6	230	-0.0068	-7470A	PPB	PPB	19 Apr 2021 13:06:19	U	SMPL
PB135792BL - 1	PBW	459	0.0487	-7470A	PPB	PPB	19 Apr 2021 13:08:34	U	SMPL
PB135792BS - 1	LCSW	17262	4.1201	-7470A	PPB	PPB	19 Apr 2021 13:10:50	U	SMPL
M2043-02 - 1	FS-WC-WATER	494	0.0571	-7470A	PPB	PPB	19 Apr 2021 13:13:06	U	SMPL
M2043-02DUP - 1	FS-WC-WATERDUP	531	0.0661	-7470A	PPB	PPB	19 Apr 2021 13:15:23	U	SMPL
M2043-02MS - 1	FS-WC-WATERMS	19025	4.5473	-7470A	PPB	PPB	19 Apr 2021 13:17:40	U	SMPL
M2043-02MSD - 1	FS-WC-WATERMSD	18784	4.4889	-7470A	PPB	PPB	19 Apr 2021 13:19:56	U	SMPL
CCV36 - 1	CCV36	20724	4.9589	-7470A	PPB	PPB	19 Apr 2021 13:22:11	U	SMPL
CCB36 - 1	CCB36	61	-0.0478	-7470A	PPB	PPB	19 Apr 2021 13:24:27	U	SMPL
PB135793BL - 1	PBW	185	-0.0177	-7470A	PPB	PPB	19 Apr 2021 13:29:17	U	SMPL
PB135793BS - 1	LCSW	15842	3.7760	-7470A	PPB	PPB	19 Apr 2021 13:31:32	U	SMPL
M2053-03 - 1	COMPOSITE-1	7396	1.7295	-7470A	PPB	PPB	19 Apr 2021 13:38:44	U	SMPL
M2053-03DUP - 1	COMPOSITE-1DUP	6750	1.5730	-7470A	PPB	PPB	19 Apr 2021 13:40:59	U	SMPL
M2053-03MS - 1	COMPOSITE-1MS	25283	6.0636	-7470A	PPB	PPB	19 Apr 2021 13:43:15	U	SMPL
M2053-03MSD - 1	COMPOSITE-1MSD	24379	5.8446	-7470A	PPB	PPB	19 Apr 2021 13:45:31	U	SMPL
PB135790TB - 1	PB135767TB	163	-0.0231	-7470A	PPB	PPB	19 Apr 2021 13:47:47	U	SMPL
PB135792TB - 1	PB135791TB	220	-0.0093	-7470A	PPB	PPB	19 Apr 2021 13:50:03	U	SMPL
CCV37 - 1	CCV37	19710	4.7132	-7470A	PPB	PPB	19 Apr 2021 13:52:20	U	SMPL
CCB37 - 1	CCB37	15	-0.0589	-7470A	PPB	PPB	19 Apr 2021 13:54:36	U	SMPL
M2048-02LX5 - 1		86	-0.0417	-7470A	PPB	PPB	19 Apr 2021 13:56:51	U	SMPL
M2048-02A - 1		16502	3.9359	-7470A	PPB	PPB	19 Apr 2021 13:59:07	U	SMPL
M2043-02LX5 - 1		71	-0.0454	-7470A	PPB	PPB	19 Apr 2021 14:01:22	U	SMPL
M2043-02A - 1		16424	3.9170	-7470A	PPB	PPB	19 Apr 2021 14:03:38	U	SMPL
M2053-03LX5 - 1		2010	0.4245	-7470A	PPB	PPB	19 Apr 2021 14:08:59	U	SMPL
M2053-03A - 1		25801	6.1891	-7470A	PPB	PPB	19 Apr 2021 14:11:14	U	SMPL
CCV38 - 1	CCV38	20643	4.9393	-7470A	PPB	PPB	19 Apr 2021 14:13:30	U	SMPL
CCB38 - 1	CCB38	107	-0.0366	-7470A	PPB	PPB	19 Apr 2021 14:15:46	U	SMPL