



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 04/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 04/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114039

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M2082-01	A0AA6	1	1.15	9.88	6.98	66.8	
M2082-02	A4302	2	1.15	9.63	8.99	92.5	
M2082-03	A4303	3	1.1	9.86	8.76	87.4	
M2082-04	A4304	4	1.16	9.54	8.84	91.6	
M2082-05	A4305	5	1.15	9.74	8.02	80.0	
M2082-06	A4306	6	1.18	9.97	8.6	84.4	
M2082-07	A4307	7	1.18	9.52	7.36	74.1	
M2082-08	A4307MS	8	1.18	9.52	7.36	74.1	
M2082-09	A4307MSD	9	1.18	9.52	7.36	74.1	
M2082-10	A4308	10	1.15	9.66	6.77	66.0	
M2082-11	A4309	11	1.11	9.51	7.15	71.9	
M2082-12	A4310	12	1.14	9.87	7.18	69.2	
M2082-13	A4311	13	1.18	9.56	6.93	68.6	
M2082-14	A4312	14	1.18	9.51	6.58	64.8	
M2082-15	A4313	15	1.18	9.66	7.09	69.7	
M2082-16	A4314	16	1.15	9.59	6.9	68.1	
M2082-17	A4315	17	1.1	9.71	7.32	72.2	
M2082-18	A4316	18	1.15	9.65	7.66	76.6	
M2082-19	A4317	19	1.15	9.94	6.94	65.9	
M2082-20	A4318	20	1.15	9.64	7.62	76.2	
M2082-21	A4319	21	1.16	9.97	5.38	47.9	
M2082-22	A4320	22	1.16	9.93	6.69	63.1	
M2082-23	A4346	23	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$