



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/27/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 04/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 04/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114041

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M2084-01	A0AA1	1	1.14	9.95	8.35	81.8	
M2084-02	A0AA2	2	1.1	9.77	8.2	81.9	
M2084-03	A0AA3	3	1.14	9.78	7.31	71.4	
M2084-04	A0AA4	4	1.12	9.54	6.76	67.0	
M2084-05	A0AA5	5	1.13	9.56	8.16	83.4	
M2084-06	A0AA7	6	1.15	9.96	6.7	63.0	
M2084-07	A0AA8	7	1.17	9.97	7.09	67.3	
M2084-08	A0AA9	8	1.12	9.88	6.73	64.0	
M2084-09	A0AB0	9	1.15	9.92	7.6	73.5	
M2084-10	A0AB1	10	1.18	9.64	7.45	74.1	
M2084-11	A0AB2	11	1.11	9.56	7.67	77.6	
M2084-12	A0AB3	12	1.16	9.81	7.94	78.4	
M2084-14	A0AB5	13	1.14	9.94	7.8	75.7	
M2084-15	A4341	14	1.12	9.64	7.39	73.6	
M2084-16	A4342	15	1.16	9.97	7.89	76.4	
M2084-17	A4343	16	1.14	9.5	6.85	68.3	
M2084-18	A4343MS	17	1.14	9.5	6.85	68.3	
M2084-19	A4343MSD	18	1.14	9.5	6.85	68.3	
M2084-20	A4345	19	1.18	9.68	7.04	68.9	
M2084-21	A4348	20	1.00	2.00	2.00	100.0	P.E. SAMPLE
M2084-22	A4344	21	1.13	9.93	6.21	57.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$