



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114045

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/19/2021	12:50
2	M2069-01		1	15	100.00	>212.0	>212.0	04/19/2021	13:20
3	M2069-01DUP		1	15	100.00	>212.0	>212.0	04/19/2021	13:45
4	M2078-01		1	14	100.00	>212.0	>212.0	04/19/2021	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114045

WorkList Name : fp41921 WorkList ID : 148145 Department : Wet-Chemistry Date : 04-19-2021 09:11:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/23/2021	Water	M2069-01	Flash Point	Cool 4 deg C	TULL02	P12	MH20A-BNM-1-D-WATEF	04/16/2021	1010A
04/23/2021	Water	M2078-01	Flash Point	Cool 4 deg C	LIRO01	P31	TWP-01	04/16/2021	1010A

Date/Time 04/19/21 12:30
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)

Date/Time 04/19/21 14:15
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)