



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 04/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:06
Out Date: 04/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114048

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M2086-01	MA4321	1	1.18	9.6	5.85	55.5	
M2086-02	MA4322	2	1.16	9.9	6.12	56.8	
M2086-03	MA4323	3	1.16	9.53	7.09	70.8	
M2086-04	MA4324	4	1.19	9.71	6.86	66.5	
M2086-05	MA4325	5	1.1	9.82	6.42	61.0	
M2086-06	MA4326	6	1.18	9.61	7.8	78.5	
M2086-07	MA4327	7	1.19	9.57	7.56	76.0	
M2086-08	MA4328	8	1.17	9.9	7.22	69.3	
M2086-09	MA4329	9	1.14	9.84	6.69	63.8	
M2086-10	MA4330	10	1.14	9.98	4.81	41.5	
M2086-11	MA4331	11	1.17	9.84	6.15	57.4	
M2086-12	MA4332	12	1.14	9.95	5.46	49.0	
M2086-13	MA4333	13	1.12	9.61	5.94	56.8	
M2086-14	MA4333D	14	1.12	9.61	5.94	56.8	
M2086-15	MA4333S	15	1.12	9.61	5.94	56.8	
M2086-16	MA4334	16	1.14	9.72	6.27	59.8	
M2086-17	MA4335	17	1.15	9.98	4.47	37.6	
M2086-18	MA4336	18	1.16	9.64	6.64	64.6	
M2086-19	MA4337	19	1.13	9.78	6.61	63.4	
M2086-20	MA4338	20	1.1	9.67	7.42	73.7	
M2086-21	MA4339	21	1.15	9.67	6.83	66.7	
M2086-22	MA4352	22	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$