



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 04/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 04/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114050

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2087-01	MA0AA1	1	1.14	9.95	8.35	81.8	
M2087-02	MA0AA2	2	1.1	9.77	8.2	81.9	
M2087-03	MA0AA3	3	1.14	9.78	7.31	71.4	
M2087-04	MA0AA4	4	1.12	9.54	6.76	67.0	
M2087-05	MA0AA5	5	1.13	9.56	8.16	83.4	
M2087-06	MA0AA7	6	1.15	9.96	6.7	63.0	
M2087-07	MA0AA8	7	1.17	9.97	7.09	67.3	
M2087-08	MA0AA9	8	1.12	9.88	6.73	64.0	
M2087-09	MA0AB0	9	1.15	9.92	7.6	73.5	
M2087-10	MA0AB1	10	1.18	9.64	7.45	74.1	
M2087-11	MA0AB2	11	1.11	9.56	7.67	77.6	
M2087-12	MA0AB3	12	1.16	9.81	7.94	78.4	
M2087-14	MA0AB5	13	1.14	9.94	7.8	75.7	
M2087-15	MA4341	14	1.12	9.64	7.39	73.6	
M2087-16	MA4342	15	1.16	9.97	7.89	76.4	
M2087-17	MA4343	16	1.14	9.5	6.85	68.3	
M2087-18	MA4343D	17	1.14	9.5	6.85	68.3	
M2087-19	MA4343S	18	1.14	9.5	6.85	68.3	
M2087-20	MA4345	19	1.18	9.68	7.04	68.9	
M2087-21	MA4340	20	1.15	9.93	7.07	67.4	
M2087-22	MA4344	21	1.13	9.93	6.21	57.7	
M2087-23	MA4351	22	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$