


LB114051

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8430e-004

C= -2.9436e-002

Rho= 0.9998746

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	163	0.000	163					1
0.2	0.200	0.226	0.026	1385	0.0 %	1385					13
2.5	2.500	2.476	-0.024	13595	0.0 %	13595					1
5.0	5.000	4.932	-0.068	26922	0.0 %	26922					1
7.5	7.500	7.612	0.112	41461	0.0 %	41461					1
10.0	10.000	9.953	-0.047	54167	0.0 %	54167					0

LB114051 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Apr 2021 14:27:13					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	163	-	0.0000	7471B	PPB	19 Apr 2021 14:31:24	S	Std
0.2 - 1	50.2	1385	-	0.2000	7471B	PPB	19 Apr 2021 14:33:41	S	Std
2.5 - 1	52.5	13595	-	2.5000	7471B	PPB	19 Apr 2021 14:35:59	S	Std
5.0 - 1	55	26922	-	5.0000	7471B	PPB	19 Apr 2021 14:38:15	S	Std
7.5 - 1	57.5	41461	-	7.5000	7471B	PPB	19 Apr 2021 14:40:32	S	Std
10.0 - 1	510	54167	-	10.0000	7471B	PPB	19 Apr 2021 14:45:21	S	Std
ICV93 - 1	ICV93	22382	4.0955	-7471B	PPB	PPB	19 Apr 2021 14:50:34	U	SMPL
ICB93 - 1	ICB93	336	0.0325	-7471B	PPB	PPB	19 Apr 2021 14:54:50	U	SMPL
CCV39 - 1	CCV39	28959	5.3077	-7471B	PPB	PPB	19 Apr 2021 14:57:05	U	SMPL
CCB39 - 1	CCB39	38	-0.0224	-7471B	PPB	PPB	19 Apr 2021 14:59:20	U	SMPL
CRA - 1	CRA	1395	0.2277	-7471B	PPB	PPB	19 Apr 2021 15:01:36	U	SMPL
HighStd - 1	HighStd	54919	10.0921	-7471B	PPB	PPB	19 Apr 2021 15:03:51	U	SMPL
ChkStd - 1	ChkStd	38663	7.0961	-7471B	PPB	PPB	19 Apr 2021 15:06:06	U	SMPL
PB135831BL - 1	PBS	-19	-0.0329	-7471B	PPB	PPB	19 Apr 2021 15:08:24	U	SMPL
PB135831BS - 1	LCSS	20373	3.7253	-7471B	PPB	PPB	19 Apr 2021 15:10:41	U	SMPL
M2063-03 - 1	TP1-2-4	3927	0.6943	-7471B	PPB	PPB	19 Apr 2021 15:22:02	U	SMPL
CCV40 - 1	CCV40	28378	5.2006	-7471B	PPB	PPB	19 Apr 2021 15:24:17	U	SMPL
CCB40 - 1	CCB40	-57	-0.0399	-7471B	PPB	PPB	19 Apr 2021 15:26:33	U	SMPL
M2063-05 - 1	TP2-0-2	44830	8.2327	-7471B	PPB	PPB	19 Apr 2021 15:28:49	U	SMPL
M2063-07 - 1	TP2-2-4	27586	5.0546	-7471B	PPB	PPB	19 Apr 2021 15:31:04	U	SMPL
M2063-09 - 1	TP3-0-2	3169	0.5546	-7471B	PPB	PPB	19 Apr 2021 15:33:21	U	SMPL
M2063-11 - 1	TP3-2-4	2919	0.5085	-7471B	PPB	PPB	19 Apr 2021 15:35:37	U	SMPL
M2070-01 - 1	JGD-3150-D3550-A	4664	0.8301	-7471B	PPB	PPB	19 Apr 2021 15:49:50	U	SMPL
M2070-05 - 1	JGD-3150-D3550-B	4396	0.7807	-7471B	PPB	PPB	19 Apr 2021 15:52:06	U	SMPL
M2075-01 - 1	TP-1-S	234	0.0137	-7471B	PPB	PPB	19 Apr 2021 15:54:22	U	SMPL
M2075-04 - 1	TP-1-D	404	0.0450	-7471B	PPB	PPB	19 Apr 2021 15:56:39	U	SMPL
M2075-07 - 1	TP-2-S	1690	0.2820	-7471B	PPB	PPB	19 Apr 2021 15:58:56	U	SMPL
M2075-10 - 1	TP-2-D	315	0.0286	-7471B	PPB	PPB	19 Apr 2021 16:01:11	U	SMPL
CCV41 - 1	CCV41	28921	5.3007	-7471B	PPB	PPB	19 Apr 2021 16:03:27	U	SMPL
CCB41 - 1	CCB41	-39	-0.0366	-7471B	PPB	PPB	19 Apr 2021 16:05:42	U	SMPL
M2075-13 - 1	TP-4-S	2350	0.4037	-7471B	PPB	PPB	19 Apr 2021 16:07:58	U	SMPL
M2075-16 - 1	TP-4-D	655	0.0913	-7471B	PPB	PPB	19 Apr 2021 16:10:13	U	SMPL
M2075-19 - 1	TP-6-S	7134	1.2853	-7471B	PPB	PPB	19 Apr 2021 16:12:29	U	SMPL
M2075-22 - 1	TP-6-D	754	0.1095	-7471B	PPB	PPB	19 Apr 2021 16:14:45	U	SMPL
M2075-25 - 1	TP-8-S	532	0.0686	-7471B	PPB	PPB	19 Apr 2021 16:17:00	U	SMPL
M2075-28 - 1	TP-8-D	483	0.0596	-7471B	PPB	PPB	19 Apr 2021 16:19:17	U	SMPL
PB135832BL - 1	PBS	319	0.0294	-7471B	PPB	PPB	19 Apr 2021 16:21:33	U	SMPL
PB135832BS - 1	LCSS	20648	3.7760	-7471B	PPB	PPB	19 Apr 2021 16:23:50	U	SMPL
M2074-01 - 1	VNPH2-B-1	24444	4.4756	-7471B	PPB	PPB	19 Apr 2021 16:26:07	U	SMPL
M2074-02 - 1	VNPH2-B-1D	20984	3.8379	-7471B	PPB	PPB	19 Apr 2021 16:28:23	U	SMPL
CCV42 - 1	CCV42	27484	5.0358	-7471B	PPB	PPB	19 Apr 2021 16:30:38	U	SMPL
CCB42 - 1	CCB42	-33	-0.0355	-7471B	PPB	PPB	19 Apr 2021 16:32:54	U	SMPL
M2074-03 - 1	VNPH2-B-2	4825	0.8598	-7471B	PPB	PPB	19 Apr 2021 16:35:09	U	SMPL
M2074-04 - 1	VNPH2-B-2D	35554	6.5231	-7471B	PPB	PPB	19 Apr 2021 16:37:25	U	SMPL
M2074-05 - 1	VNPH2-B-3	6800	1.2238	-7471B	PPB	PPB	19 Apr 2021 16:39:40	U	SMPL
M2074-06 - 1	VNPH2-B-3D	28760	5.2710	-7471B	PPB	PPB	19 Apr 2021 16:41:57	U	SMPL
M2074-07 - 1	VNPH2-B-4	14415	2.6272	-7471B	PPB	PPB	19 Apr 2021 16:44:13	U	SMPL
M2074-08 - 1	VNPH2-B-5	12486	2.2717	-7471B	PPB	PPB	19 Apr 2021 16:46:29	U	SMPL
M2074-09 - 1	VNPH2-B-6	11130	2.0218	-7471B	PPB	PPB	19 Apr 2021 16:48:45	U	SMPL
M2074-10 - 1	VNPH2-B-6D	12396	2.2551	-7471B	PPB	PPB	19 Apr 2021 16:51:02	U	SMPL
M2074-11 - 1	VNPH2-B-7	9966	1.8073	-7471B	PPB	PPB	19 Apr 2021 16:53:18	U	SMPL
M2074-12 - 1	VNPH2-B-8	16034	2.9256	-7471B	PPB	PPB	19 Apr 2021 16:55:34	U	SMPL
CCV43 - 1	CCV43	27819	5.0976	-7471B	PPB	PPB	19 Apr 2021 16:57:49	U	SMPL
CCB43 - 1	CCB43	-36	-0.0361	-7471B	PPB	PPB	19 Apr 2021 17:00:05	U	SMPL
M2074-13 - 1	VNPH2-B-9	7405	1.3353	-7471B	PPB	PPB	19 Apr 2021 17:02:20	U	SMPL
M2074-14 - 1	VNPH2-B-10	6401	1.1503	-7471B	PPB	PPB	19 Apr 2021 17:04:36	U	SMPL
M2074-15 - 1	VNPH2-B-10D	27325	5.0065	-7471B	PPB	PPB	19 Apr 2021 17:06:51	U	SMPL
M2074-16 - 1	VNPH2-B-11	17136	3.1287	-7471B	PPB	PPB	19 Apr 2021 17:09:07	U	SMPL
PB135833BL - 1	PBS	53	-0.0197	-7471B	PPB	PPB	19 Apr 2021 17:22:58	U	SMPL
PB135833BS - 1	LCSS	22245	4.0703	-7471B	PPB	PPB	19 Apr 2021 17:25:14	U	SMPL
CCV44 - 1	CCV44	28076	5.1449	-7471B	PPB	PPB	19 Apr 2021 17:27:30	U	SMPL
CCB44 - 1	CCB44	-24	-0.0339	-7471B	PPB	PPB	19 Apr 2021 17:29:46	U	SMPL
M2076-03 - 1	SB-13-14.5-15.0	150	-0.0018	-7471B	PPB	PPB	19 Apr 2021 17:41:07	U	SMPL
M2076-04 - 1	SB-14-0-2.0	1913	0.3231	-7471B	PPB	PPB	19 Apr 2021 17:43:23	U	SMPL
M2076-06 - 1	SB-14-14.5-15.0	2342	0.4022	-7471B	PPB	PPB	19 Apr 2021 17:45:40	U	SMPL
M2076-07 - 1	SB-15-0-2.0	4406	0.7826	-7471B	PPB	PPB	19 Apr 2021 17:47:56	U	SMPL
M2076-09 - 1	SB-15-9.5-10.0	2284	0.3915	-7471B	PPB	PPB	19 Apr 2021 17:50:13	U	SMPL
M2076-10 - 1	SB-16-0-2.0	7967	1.4389	-7471B	PPB	PPB	19 Apr 2021 17:52:29	U	SMPL
CCV45 - 1	CCV45	29008	5.3167	-7471B	PPB	PPB	19 Apr 2021 17:54:45	U	SMPL
CCB45 - 1	CCB45	-34	-0.0357	-7471B	PPB	PPB	19 Apr 2021 17:57:01	U	SMPL
M2076-12 - 1	SB-16-9.5-10.0	1777	0.2981	-7471B	PPB	PPB	19 Apr 2021 17:59:17	U	SMPL
M2076-13 - 1	SB-17-0-2.0	2478	0.4273	-7471B	PPB	PPB	19 Apr 2021 18:01:34	U	SMPL
M2076-15 - 1	SB-17-14.5-15.0	3830	0.6764	-7471B	PPB	PPB	19 Apr 2021 18:03:50	U	SMPL
M2076-16 - 1	SB-18-0-2.0	2699	0.4680	-7471B	PPB	PPB	19 Apr 2021 18:06:05	U	SMPL
M2076-18 - 1	SB-18-14.5-15.0	9764	1.7701	-7471B	PPB	PPB	19 Apr 2021 18:08:22	U	SMPL
M2076-19 - 1	SB-19-0-2.0	2619	0.4532	-7471B	PPB	PPB	19 Apr 2021 18:10:38	U	SMPL
M2077-01 - 1	SB-19-14.5-15.0	33433	6.1322	-7471B	PPB	PPB	19 Apr 2021 18:12:54	U	SMPL
M2077-02 - 1	SB-20-0-2.0	6457	1.1606	-7471B	PPB	PPB	19 Apr 2021 18:15:11	U	SMPL
CCV46 - 1	CCV46	29152	5.3432	-7471B	PPB	PPB	19 Apr 2021 18:22:00	U	SMPL
CCB46 - 1	CCB46	-46	-0.0379	-7471B	PPB	PPB	19 Apr 2021 18:24:18	U	SMPL