



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114053

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/19/2021	14:25
2	CAL2	1	Water	NA	NA	20.3	7.00	04/19/2021	14:26
3	CAL3	1	Water	NA	NA	20.3	10.03	04/19/2021	14:27
4	ICV	1	Water	NA	NA	20.4	6.99	04/19/2021	14:30
5	CCV1	1	Water	NA	NA	20.3	2.01	04/19/2021	14:32
6	M2090-03	1	Solid	20.03	20	21.5	11.06	04/19/2021	14:35
7	M2090-03DUP	1	Solid	20.04	20	21.6	11.07	04/19/2021	14:36
8	M2090-08	1	Solid	20.02	20	21.0	11.02	04/19/2021	14:38
9	M2090-13	1	Solid	20.03	20	20.8	10.97	04/19/2021	14:40
10	M2090-18	1	Solid	20.02	20	20.9	11.46	04/19/2021	14:42
11	M2090-22	1	Solid	20.04	20	20.8	10.30	04/19/2021	14:46
12	CCV2	1	Water	NA	NA	20.4	12.02	04/19/2021	14:48

WORKLIST(Hardcopy Internal Chain)

18714053

WorkList Name : corrosivity m2090

WorkList ID : 148167

Department : Wet-Chemistry

Date : 04-19-2021 14:12:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/16/2021	Solid	M2090-03	pH	Cool 4 deg C	VERI01	O51	SEDGWICK-COMP-1	04/16/2021	9045D
04/16/2021	Solid	M2090-08	pH	Cool 4 deg C	VERI01	O51	SEDGWICK-COMP-2	04/16/2021	9045D
04/16/2021	Solid	M2090-13	pH	Cool 4 deg C	VERI01	O51	SEDGWICK-COMP-3	04/16/2021	9045D
04/16/2021	Solid	M2090-18	pH	Cool 4 deg C	VERI01	O51	SEDGWICK-COMP-4	04/16/2021	9045D
04/16/2021	Solid	M2090-22	pH	Cool 4 deg C	VERI01	O51	SEDGWICK-COMP-5	04/16/2021	9045D

Date/Time 04-19-21 14:15
 Raw Sample Received by: JP (wet room)
 Raw Sample Relinquished by: JP (sm1)

Date/Time 04-19-21 15:30
 Raw Sample Received by: JP (sm1)
 Raw Sample Relinquished by: JP (wet room)