

LB 114059


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54816



A= 0.0000e+000

B= 1.8196e-004

C= 4.5522e-003

Rho= 0.9999668

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.018	0.018	72	0.000	72					-
0.2	0.200	0.202	0.002	1084	0.0 %	1084					-
2.5	2.500	2.477	-0.023	13589	0.0 %	13589					-
5.0	5.000	4.968	-0.032	27280	0.0 %	27280					-
7.5	7.500	7.556	0.056	41501	0.0 %	41501					-
10.0	10.000	9.979	-0.021	54816	0.0 %	54816					0

LB114059 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Apr 2021 10:42:04

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	72	-	0.00007471B	PPB	20 Apr 2021 10:57:00	S	Std
0.2 - 1	50.2	1084	-	0.20007471B	PPB	20 Apr 2021 10:59:16	S	Std
2.5 - 1	52.5	13589	-	2.50007471B	PPB	20 Apr 2021 11:01:33	S	Std
5.0 - 1	55	27280	-	5.00007471B	PPB	20 Apr 2021 11:03:50	S	Std
7.5 - 1	57.5	41501	-	7.50007471B	PPB	20 Apr 2021 11:06:06	S	Std
10.0 - 1	510	54816	-	10.00007471B	PPB	20 Apr 2021 11:08:23	S	Std
ICV94 - 1	ICV94	20826	3.7940	-7471B	PPB	20 Apr 2021 11:11:10	U	SMPL
ICB94 - 1	ICB94	-150	-0.0227	-7471B	PPB	20 Apr 2021 11:13:25	U	SMPL
CCV47 - 1	CCV47	27812	5.0652	-7471B	PPB	20 Apr 2021 11:15:40	U	SMPL
CCB47 - 1	CCB47	-174	-0.0271	-7471B	PPB	20 Apr 2021 11:17:55	U	SMPL
CRA - 1	CRA	1048	0.1952	-7471B	PPB	20 Apr 2021 11:20:10	U	SMPL
HighStd - 1	HighStd	54606	9.9407	-7471B	PPB	20 Apr 2021 11:22:25	U	SMPL
ChkStd - 1	ChkStd	38451	7.0011	-7471B	PPB	20 Apr 2021 11:24:42	U	SMPL
M2063-01 - 1	TP1-0-2	5680	1.0381	-7471B	PPB	20 Apr 2021 11:26:59	U	SMPL
M2063-01DUP - 1	TP1-0-2DUP	5323	0.9731	-7471B	PPB	20 Apr 2021 11:29:16	U	SMPL
M2063-01MS - 1	TP1-0-2MS	25724	4.6853	-7471B	PPB	20 Apr 2021 11:31:32	U	SMPL
M2063-01MSD - 1	TP1-0-2MSD	24466	4.4564	-7471B	PPB	20 Apr 2021 11:33:49	U	SMPL
M2074-17 - 1	VNPH2-B-12	13810	2.5174	-7471B	PPB	20 Apr 2021 11:36:06	U	SMPL
M2074-17DUP - 1	VNPH2-B-12DUP	13707	2.4987	-7471B	PPB	20 Apr 2021 11:38:22	U	SMPL
M2074-17MS - 1	VNPH2-B-12MS	37219	6.7769	-7471B	PPB	20 Apr 2021 11:40:38	U	SMPL
CCV48 - 1	CCV48	28246	5.1442	-7471B	PPB	20 Apr 2021 11:42:53	U	SMPL
CCB48 - 1	CCB48	-47	-0.0040	-7471B	PPB	20 Apr 2021 11:45:10	U	SMPL
M2074-17MSD - 1	VNPH2-B-12MSD	38141	6.9447	-7471B	PPB	20 Apr 2021 11:47:25	U	SMPL
M2076-01 - 1	SB-13-0-2.0	6221	1.1365	-7471B	PPB	20 Apr 2021 11:49:41	U	SMPL
M2076-01DUP - 1	SB-13-0-2.0DUP	6343	1.1587	-7471B	PPB	20 Apr 2021 11:51:57	U	SMPL
M2076-01MS - 1	SB-13-0-2.0MS	28550	5.1995	-7471B	PPB	20 Apr 2021 11:54:13	U	SMPL
M2076-01MSD - 1	SB-13-0-2.0MSD	29014	5.2839	-7471B	PPB	20 Apr 2021 11:56:29	U	SMPL
M2077-04 - 1	SB-20-14.5-15.0	7334	1.3390	-7471B	PPB	20 Apr 2021 11:58:45	U	SMPL
M2077-05 - 1	SB-20-14.5-15.0DUP	8891	1.6224	-7471B	PPB	20 Apr 2021 12:03:42	U	SMPL
M2077-06 - 1	SB-21-0-2.0	4039	0.7395	-7471B	PPB	20 Apr 2021 12:05:59	U	SMPL
M2077-08 - 1	SB-21-14.5-15.0	11924	2.1742	-7471B	PPB	20 Apr 2021 12:08:16	U	SMPL
CCV49 - 1	CCV49	28195	5.1349	-7471B	PPB	20 Apr 2021 12:10:31	U	SMPL
CCB49 - 1	CCB49	-25	0.0000	-7471B	PPB	20 Apr 2021 12:12:46	U	SMPL
M2063-01LX5 - 1		939	0.1754	-7471B	PPB	20 Apr 2021 12:15:02	U	SMPL
M2063-01A - 1		30710	5.5925	-7471B	PPB	20 Apr 2021 12:17:17	U	SMPL
M2074-17LX5 - 1		2717	0.4989	-7471B	PPB	20 Apr 2021 12:19:33	U	SMPL
M2074-17A - 1		37470	6.8226	-7471B	PPB	20 Apr 2021 12:21:48	U	SMPL
M2076-01LX5 - 1		1099	0.2045	-7471B	PPB	20 Apr 2021 12:24:03	U	SMPL
M2076-01A - 1		28967	5.2754	-7471B	PPB	20 Apr 2021 12:26:19	U	SMPL
CCV50 - 1	CCV50	28224	5.1402	-7471B	PPB	20 Apr 2021 12:28:36	U	SMPL
CCB50 - 1	CCB50	-37	-0.0022	-7471B	PPB	20 Apr 2021 12:30:52	U	SMPL