


LB114669

7470A

INSTRUMENT ID: CV1

 μ Abs.:

40201



A= 0.0000e+000

B= 2.4690e-004

C= -2.6179e-003

Rho= 0.9997807

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	150	0.000	150					
0.2	0.200	0.212	0.012	871	0.0 %	871					
2.5	2.500	2.409	-0.091	9767	0.0 %	9767					
5.0	5.000	4.980	-0.020	20181	0.0 %	20181					
7.5	7.500	7.641	0.141	30958	0.0 %	30958					
10.0	10.000	9.923	-0.077	40201	0.0 %	40201					

6
-4
0
2
-1

LB114067 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Apr 2021 13:11:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	150	-	0.0000	7470A	PPB	20 Apr 2021 14:23:43	S	Std
0.2 - 1	50.2	871	-	0.2000	7470A	PPB	20 Apr 2021 14:26:00	S	Std
2.5 - 1	52.5	9767	-	2.5000	7470A	PPB	20 Apr 2021 14:28:17	S	Std
5.0 - 1	55	20181	-	5.0000	7470A	PPB	20 Apr 2021 14:30:33	S	Std
7.5 - 1	57.5	30958	-	7.5000	7470A	PPB	20 Apr 2021 14:32:50	S	Std
10.0 - 1	510	40201	-	10.0000	7470A	PPB	20 Apr 2021 14:35:05	S	Std
ICV95 - 1	ICV95	14691	3.6246	-7470A	PPB	20 Apr 2021 14:37:46	U	SMPL	
ICB95 - 1	ICB95	-117	-0.0315	-7470A	PPB	20 Apr 2021 14:40:00	U	SMPL	
CCV51 - 1	CCV51	21844	5.3907	-7470A	PPB	20 Apr 2021 14:42:15	U	SMPL	
CCB51 - 1	CCB51	-129	-0.0345	-7470A	PPB	20 Apr 2021 14:44:30	U	SMPL	
CRA - 1	CRA	798	0.1944	-7470A	PPB	20 Apr 2021 14:46:45	U	SMPL	
HighStd - 1	HighStd	40029	9.8807	-7470A	PPB	20 Apr 2021 14:49:00	U	SMPL	
ChkStd - 1	ChkStd	28146	6.9467	-7470A	PPB	20 Apr 2021 14:51:14	U	SMPL	
PB135834BL - 1	PBW	-246	-0.0634	-7470A	PPB	20 Apr 2021 14:53:31	U	SMPL	
PB135834BS - 1	LCSW	14827	3.6582	-7470A	PPB	20 Apr 2021 14:55:49	U	SMPL	
M2049-03 - 1	SB-08-COMP	-159	-0.0419	-7470A	PPB	20 Apr 2021 14:58:05	U	SMPL	
M2049-07 - 1	SB-09-COMP	6	-0.0011	-7470A	PPB	20 Apr 2021 15:00:21	U	SMPL	
M2049-11 - 1	SB-10-COMP	25	0.0036	-7470A	PPB	20 Apr 2021 15:02:38	U	SMPL	
M2049-14 - 1	SB-11-COMP	52	0.0102	-7470A	PPB	20 Apr 2021 15:04:54	U	SMPL	
M2049-17 - 1	SB-12-COMP	68	0.0142	-7470A	PPB	20 Apr 2021 15:07:09	U	SMPL	
CCV52 - 1	CCV52	21534	5.3142	-7470A	PPB	20 Apr 2021 15:09:25	U	SMPL	
CCB52 - 1	CCB52	-114	-0.0308	-7470A	PPB	20 Apr 2021 15:11:40	U	SMPL	
M2070-02 - 1	JGD-3150-D3550-A	1	-0.0024	-7470A	PPB	20 Apr 2021 15:13:56	U	SMPL	
M2070-06 - 1	JGD-3150-D3550-B	34	0.0058	-7470A	PPB	20 Apr 2021 15:16:11	U	SMPL	
M2076-02 - 1	SB-13-COMP	50	0.0097	-7470A	PPB	20 Apr 2021 15:18:27	U	SMPL	
M2076-05 - 1	SB-14-COMP	106	0.0236	-7470A	PPB	20 Apr 2021 15:20:42	U	SMPL	
M2076-08 - 1	SB-15-COMP	99	0.0218	-7470A	PPB	20 Apr 2021 15:22:57	U	SMPL	
M2076-11 - 1	SB-16-COMP	47	0.0090	-7470A	PPB	20 Apr 2021 15:25:13	U	SMPL	
M2076-14 - 1	SB-17-COMP	50	0.0097	-7470A	PPB	20 Apr 2021 15:27:30	U	SMPL	
M2076-17 - 1	SB-18-COMP	63	0.0129	-7470A	PPB	20 Apr 2021 15:29:46	U	SMPL	
M2076-20 - 1	SB-19-COMP	81	0.0174	-7470A	PPB	20 Apr 2021 15:32:04	U	SMPL	
M2077-03 - 1	SB-20-COMP	100	0.0221	-7470A	PPB	20 Apr 2021 15:34:19	U	SMPL	
CCV53 - 1	CCV53	21617	5.3347	-7470A	PPB	20 Apr 2021 15:36:35	U	SMPL	
CCB53 - 1	CCB53	-92	-0.0253	-7470A	PPB	20 Apr 2021 15:38:51	U	SMPL	
M2077-07 - 1	SB-21-COMP	85	0.0184	-7470A	PPB	20 Apr 2021 15:41:08	U	SMPL	
M2077-07DUP - 1	SB-21-COMPDUP	87	0.0189	-7470A	PPB	20 Apr 2021 15:43:23	U	SMPL	
M2077-07MS - 1	SB-21-COMPMS	15986	3.9444	-7470A	PPB	20 Apr 2021 15:45:38	U	SMPL	
M2077-07MSD - 1	SB-21-COMPMSD	13965	3.4454	-7470A	PPB	20 Apr 2021 15:47:54	U	SMPL	
PB135835BL - 1	PBW	-104	-0.0283	-7470A	PPB	20 Apr 2021 15:50:10	U	SMPL	
PB135835BS - 1	LCSW	13719	3.3846	-7470A	PPB	20 Apr 2021 15:52:26	U	SMPL	
M2075-02 - 1	TP-1-S	-136	-0.0362	-7470A	PPB	20 Apr 2021 15:54:41	U	SMPL	
M2075-02DUP - 1	TP-1-SDUP	-20	-0.0076	-7470A	PPB	20 Apr 2021 15:56:58	U	SMPL	
M2075-02MS - 1	TP-1-SMS	14405	3.5540	-7470A	PPB	20 Apr 2021 15:59:15	U	SMPL	
M2075-02MSD - 1	TP-1-SMSD	14031	3.4617	-7470A	PPB	20 Apr 2021 16:01:30	U	SMPL	
CCV54 - 1	CCV54	19129	4.7204	-7470A	PPB	20 Apr 2021 16:03:46	U	SMPL	
CCB54 - 1	CCB54	-148	-0.0392	-7470A	PPB	20 Apr 2021 16:06:01	U	SMPL	
M2075-05 - 1	TP-1-D	-31	-0.0103	-7470A	PPB	20 Apr 2021 16:08:18	U	SMPL	
M2075-08 - 1	TP-2-S	8	-0.0006	-7470A	PPB	20 Apr 2021 16:10:33	U	SMPL	
M2075-11 - 1	TP-2-D	20	0.0023	-7470A	PPB	20 Apr 2021 16:12:48	U	SMPL	
M2075-14 - 1	TP-4-S	37	0.0065	-7470A	PPB	20 Apr 2021 16:15:04	U	SMPL	
M2075-17 - 1	TP-4-D	46	0.0087	-7470A	PPB	20 Apr 2021 16:17:20	U	SMPL	
M2075-20 - 1	TP-6-S	51	0.0100	-7470A	PPB	20 Apr 2021 16:19:36	U	SMPL	
M2075-23 - 1	TP-6-D	42	0.0078	-7470A	PPB	20 Apr 2021 16:21:52	U	SMPL	
M2075-26 - 1	TP-8-S	54	0.0107	-7470A	PPB	20 Apr 2021 16:24:09	U	SMPL	
M2075-29 - 1	TP-8-D	58	0.0117	-7470A	PPB	20 Apr 2021 16:26:26	U	SMPL	
CCV55 - 1	CCV55	19990	4.9330	-7470A	PPB	20 Apr 2021 16:28:42	U	SMPL	
CCB55 - 1	CCB55	-65	-0.0187	-7470A	PPB	20 Apr 2021 16:30:58	U	SMPL	
PB135834TB - 1	PB135786TB	26	0.0038	-7470A	PPB	20 Apr 2021 16:33:14	U	SMPL	
PB135835TB - 1	PB135813TB	36	0.0063	-7470A	PPB	20 Apr 2021 16:35:29	U	SMPL	
M2077-07LX5 - 1		99	0.0218	-7470A	PPB	20 Apr 2021 16:37:44	U	SMPL	
M2077-07A - 1		15623	3.8547	-7470A	PPB	20 Apr 2021 16:39:59	U	SMPL	
M2075-02LX5 - 1		-2	-0.0031	-7470A	PPB	20 Apr 2021 16:42:15	U	SMPL	
M2075-02A - 1		16001	3.9481	-7470A	PPB	20 Apr 2021 16:44:31	U	SMPL	
CCV56 - 1	CCV56	21639	5.3401	-7470A	PPB	20 Apr 2021 16:46:47	U	SMPL	
CCB56 - 1	CCB56	-149	-0.0394	-7470A	PPB	20 Apr 2021 16:49:03	U	SMPL	