



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114076

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/20/2021	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	04/20/2021	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	04/20/2021	15:02
4	ICV	1	Water	NA	NA	20.3	6.99	04/20/2021	15:05
5	CCV1	1	Water	NA	NA	20.3	2.01	04/20/2021	15:08
6	M2095-02	1	Solid	20.02	20	20.2	6.14	04/20/2021	16:50
7	M2106-02	1	Solid	20.03	20	21.8	7.30	04/20/2021	16:52
8	M2106-05	1	Solid	20.01	20	22.3	7.19	04/20/2021	17:00
9	M2106-08	1	Solid	20.03	20	21.8	7.74	04/20/2021	17:02
10	M2106-11	1	Solid	20.04	20	21.6	8.31	04/20/2021	17:05
11	M2106-14	1	Solid	20.04	20	21.7	8.18	04/20/2021	17:06
12	M2106-17	1	Solid	20.03	20	21.8	8.10	04/20/2021	17:10
13	M2106-20	1	Solid	20.04	20	21.8	7.98	04/20/2021	17:13
14	M2107-02	1	Solid	20.01	20	21.6	7.80	04/20/2021	17:15
15	M2107-05	1	Solid	20.03	20	21.5	7.61	04/20/2021	17:18
16	CCV2	1	Water	NA	NA	21.2	12.02	04/20/2021	17:20
17	M2107-09	1	Solid	20.04	20	21.5	8.65	04/20/2021	17:22
18	M2107-12	1	Solid	20.03	20	21.4	8.33	04/20/2021	17:25
19	M2107-15	1	Solid	20.04	20	21.9	8.08	04/20/2021	17:28
20	M2107-18	1	Solid	20.03	20	21.5	8.07	04/20/2021	17:30

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M2107-21	1	Solid	20.04	20	21.7	8.41	04/20/2021	17:31
22	M2107-21DUP	1	Solid	20.03	20	21.8	8.42	04/20/2021	17:32
23	CCV3	1	Water	NA	NA	20.4	2.01	04/20/2021	17:35

WORKLIST(Hardcopy Internal Chain)

14076

WorkList Name : corrosivity m2106

WorkList ID : 148201

Department : Wet-Chemistry

Date : 04-20-2021 16:42:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/19/2021	Solid	M2095-02	Corrosivity	Cool 4 deg C	FIRS02	P11	WC-1	04/19/2021	9045D
04/16/2021	Solid	M2106-02	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-22-COMP	04/16/2021	9045D
04/16/2021	Solid	M2106-05	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-23-COMP	04/16/2021	9045D
04/16/2021	Solid	M2106-08	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-24-COMP	04/16/2021	9045D
04/19/2021	Solid	M2106-11	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-25-COMP	04/19/2021	9045D
04/19/2021	Solid	M2106-14	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-26-COMP	04/19/2021	9045D
04/19/2021	Solid	M2106-17	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-27-COMP	04/19/2021	9045D
04/19/2021	Solid	M2106-20	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-28-COMP	04/19/2021	9045D
04/19/2021	Solid	M2107-02	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-29-COMP	04/19/2021	9045D
04/19/2021	Solid	M2107-05	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-30-COMP	04/19/2021	9045D
04/20/2021	Solid	M2107-09	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-31-COMP	04/20/2021	9045D
04/20/2021	Solid	M2107-12	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-32-COMP	04/20/2021	9045D
04/20/2021	Solid	M2107-15	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-33-COMP	04/20/2021	9045D
04/20/2021	Solid	M2107-18	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-34-COMP	04/20/2021	9045D
04/20/2021	Solid	M2107-21	Corrosivity	Cool 4 deg C	LIRO01	O31	SB-35-COMP	04/20/2021	9045D

Date/Time 04/20/21 16:45
 Raw Sample Received by: JS (wetchem)
 Raw Sample Relinquished by: JS (wetchem)

Date/Time 04/20/21 18:00
 Raw Sample Received by: KS (wetchem)
 Raw Sample Relinquished by: JS (wetchem)