



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB114077

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/20/2021	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	04/20/2021	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	04/20/2021	15:02
4	ICV	1	Water	NA	NA	20.3	6.99	04/20/2021	15:05
5	CCV1	1	Water	NA	NA	20.3	2.01	04/20/2021	15:08
6	M2101-04	1	Water	NA	NA	20.5	7.14	04/20/2021	15:10
7	M2101-04DUP	1	Water	NA	NA	20.6	7.15	04/20/2021	15:11
8	M2101-05	1	Water	NA	NA	20.6	6.18	04/20/2021	15:15
9	M2101-06	1	Water	NA	NA	20.3	5.21	04/20/2021	15:16
10	CCV2	1	Water	NA	NA	20.3	12.02	04/20/2021	15:19

RB 114077

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2101

WorkList ID : 148197

Department : Wet-Chemistry

Date : 04-20-2021 14:45:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2021	Water	M2101-04	pH	Cool 4 deg C	PSEG03	O32	30834	04/20/2021	9040C
04/20/2021	Water	M2101-05	pH	Cool 4 deg C	PSEG03	O32	31044	04/20/2021	9040C
04/20/2021	Water	M2101-06	pH	Cool 4 deg C	PSEG03	O32	31056	04/20/2021	9040C

Date/Time 04/20/21 14:50
 Raw Sample Received by: JP (wet chem)
 Raw Sample Relinquished by: CP SM

Date/Time 04/20/21 15:50
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: JP (wet chem)