

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
CCV1	9.5544	0.0702	0.73	TOC	
CCB1	0.3846	0.0226	5.88	TOC	
LB114081BLW	0.3455	0.0251	7.27	TOC	
LB114081BSW.....	10.6454...	0.2930..	2.75...	TOC	..
M2114-07	0.7979	0.0419	5.25	TOC	
M2114-01X10	2.1913	0.0408	1.86	TOC	
M2114-02X5.....	4.9254...	0.0614..	1.25...	TOC	..
M2114-03	5.6257	0.3464	6.16	TOC	
M2114-04	5.5862	0.2821	5.05	TOC	
M2114-05MS.....	12.9430...	0.3712..	2.87...	TOC	..
M2114-06MSD	13.3264	0.3418	2.56	TOC	
CCV2	9.6130	0.1428	1.49	TOC	
CCB2.....	0.2899...	0.0268..	9.25...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	Sample	2	2021/04/21 09:23	
TOC 0 - 20 ppmC	Sample	3	2021/04/21 09:45	
TOC 0 - 20 ppmC	Sample	4	2021/04/21 10:08	
TOC 0 - 20 ppmC	...Sample	.. 5..	2021/04/21 10:32	..
TOC 0 - 20 ppmC	Sample	6	2021/04/21 11:19	
TOC 0 - 20 ppmC	Sample	7	2021/04/21 11:42	
TOC 0 - 20 ppmC	...Sample	.. 8..	2021/04/21 12:06	..
TOC 0 - 20 ppmC	Sample	9	2021/04/21 12:31	
TOC 0 - 20 ppmC	Sample	10	2021/04/21 12:55	
TOC 0 - 20 ppmC	...Sample	.. 11..	2021/04/21 13:22	..
TOC 0 - 20 ppmC	Sample	12	2021/04/21 13:48	
TOC 0 - 20 ppmC	Sample	2	2021/04/21 14:13	
TOC 0 - 20 ppmC	...Sample	.. 3..	2021/04/21 14:45	..

```
=====
Sample ID:   CCV1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   04210838
Cal. Curve:  TOC WATER 9060A          Timestamp: 2021/04/21 09:23
Operator ID: AM AP                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	9.6070	4.8035	414098	-0.610	-0.411	126
2	9.6060	4.8030	414055	-0.537	-0.338	130
3	9.4580	4.7290	407673	-0.516	-0.316	123
4	9.5467	4.7734	411499	-0.533	-0.333	127

```
-----
<<<Statistics>>>   Mean:   9.5544   Std Dev:   0.0702   RSD: 0.73
=====
```

```
=====
Sample ID:   CCB1                      Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   04210838
Cal. Curve:  TOC WATER 9060A          Timestamp: 2021/04/21 09:45
Operator ID: AM AP                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.3981	0.1991	17160	-0.626	-0.428	94
2	0.3527	0.1764	15205	-0.599	-0.404	94
3	0.3847	0.1924	16583	-0.612	-0.414	95
4	0.4029	0.2014	17365	-0.623	-0.425	96

```
-----
<<<Statistics>>>   Mean:   0.3846   Std Dev:   0.0226   RSD: 5.88
=====
```

```
=====
Sample ID:   LB114081BLW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   04210947
Cal. Curve:  TOC WATER 9060A          Timestamp: 2021/04/21 10:08
Operator ID: AM AP                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.3327	0.1664	14341	-0.576	-0.376	94
2	0.3330	0.1665	14352	-0.538	-0.339	93
3	0.3832	0.1916	16516	-0.548	-0.350	94
4	0.3332	0.1666	14361	-0.542	-0.342	93

```
-----
<<<Statistics>>>   Mean:   0.3455   Std Dev:   0.0251   RSD: 7.27
=====
```

```
=====
Sample ID:   LB114081BSW              Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   04210947
Cal. Curve:  TOC WATER 9060A          Timestamp: 2021/04/21 10:32
Operator ID: AM AP                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	11.0732	5.5366	477297	-0.485	-0.286	153
2	10.5760	5.2880	455867	-0.383	-0.185	125
3	10.4136	5.2068	448863	-0.353	-0.156	132
4	10.5188	5.2594	453400	-0.346	-0.147	124

```
-----
<<<Statistics>>>   Mean:  10.6454   Std Dev:   0.2930   RSD: 2.75
=====
```

```
=====
Sample ID:   M2114-07                 Mode:      TOC
Method:      TOC 0 - 20 ppmC           Filename:   04211057
Cal. Curve:  TOC WATER 9060A          Timestamp: 2021/04/21 11:19
Operator ID: AM AP                     Sample Type: Sample
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning	Ending	Integration
-------	-------	------	----------	-----------	--------	-------------

<<<Statistics>>>	Mean:	0.7979	Std Dev:	0.0419	RSD: 5.25
------------------	-------	--------	----------	--------	-----------

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	2.2078	1.1039	95165	-0.070	0.130	108
2	2.2047	1.1023	95029	-0.086	0.113	107
3	2.1311	1.0655	91858	-0.064	0.134	115
4	2.2215	1.1107	95754	-0.074	0.124	114

<<<Statistics>>> Mean: 2.1913 Std Dev: 0.0408 RSD: 1.86

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	4.9845	2.4922	214850	0.012	0.209	125
2	4.9592	2.4796	213759	0.027	0.227	120
3	4.8445	2.4223	208816	0.068	0.267	139
4	4.9134	2.4567	211785	0.055	0.253	126

<<<Statistics>>> Mean: 4.9254 Std Dev: 0.0614 RSD: 1.25

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	6.1166	3.0583	263649	0.135	0.333	134
2	5.6208	2.8104	242277	0.172	0.372	132
3	5.4039	2.7020	232930	0.175	0.374	137
4	5.3615	2.6808	231101	0.176	0.375	134

<<<Statistics>>> Mean: 5.6257 Std Dev: 0.3464 RSD: 6.16

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	5.9822	2.9911	257857	0.212	0.411	134
2	5.5622	2.7811	239753	0.265	0.463	134
3	5.4785	2.7392	236143	0.289	0.489	132
4	5.3221	2.6611	229403	0.327	0.524	131

<<<Statistics>>> Mean: 5.5862 Std Dev: 0.2821 RSD: 5.05

Sample ID: M2114-05MS Mode: TOC
 Method: TOC 0 - 20 ppmC Filename: 04211057
 Cal. Curve: TOC WATER 9060A Timestamp: 2021/04/21 13:22
 Operator ID: AM AP Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.4681	6.7341	580527	0.309	0.506	169
2	12.9376	6.4688	557659	0.411	0.610	160
3	12.7192	6.3596	548247	0.434	0.632	159
4	12.6470	6.3235	545134	0.499	0.698	151

<<<Statistics>>> Mean: 12.9430 Std Dev: 0.3712 RSD: 2.87

Sample ID: M2114-06MSD Mode: TOC
 Method: TOC 0 - 20 ppmC Filename: 04211057
 Cal. Curve: TOC WATER 9060A Timestamp: 2021/04/21 13:48
 Operator ID: AM AP Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	13.7666	6.8833	593393	0.541	0.739	166
2	13.2691	6.6346	571949	0.602	0.801	148
3	13.3351	6.6676	574795	0.585	0.785	173
4	12.9348	6.4674	557540	0.626	0.825	147

<<<Statistics>>> Mean: 13.3264 Std Dev: 0.3418 RSD: 2.56

Sample ID: CCV2 Mode: TOC
 Method: TOC 0 - 20 ppmC Filename: 04211350
 Cal. Curve: TOC WATER 9060A Timestamp: 2021/04/21 14:13
 Operator ID: AM AP Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	9.6823	4.8411	417342	0.556	0.754	129
2	9.6588	4.8294	416331	0.621	0.821	124
3	9.7099	4.8549	418532	0.625	0.824	121
4	9.4011	4.7005	405221	0.613	0.811	126

<<<Statistics>>> Mean: 9.6130 Std Dev: 0.1428 RSD: 1.49

Sample ID: CCB2 Mode: TOC
 Method: TOC 0 - 20 ppmC Filename: 04211424
 Cal. Curve: TOC WATER 9060A Timestamp: 2021/04/21 14:45
 Operator ID: AM AP Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2866	0.1433	12352	0.434	0.633	90
2	0.2828	0.1414	12189	0.413	0.612	89
3	0.2632	0.1316	11344	0.421	0.618	91
4	0.3271	0.1636	14101	0.391	0.586	90

<<<Statistics>>> Mean: 0.2899 Std Dev: 0.0268 RSD: 9.25

Sample ID	Result	Std. Dev.	RSD	Mode	ALT
0.0PPM	10935	2079	19.01	TOC	
0.5PPM	30350	1328	4.38	TOC	
1.0PPM	50065	2517	5.03	TOC	
2.0PPM.....	88711...	2814..	3.17...	TOC	..
5.0PPM	202799	1586	0.78	TOC	
10.0PPM	442148	5317	1.20	TOC	
20.0PPM.....	868135...	4860..	0.56...	TOC	..
ICV	10.4223	0.2300	2.21	TOC	
ICB	0.2492	0.0428	17.18	TOC	
IC-20PPM.....	0.2729...	0.0352..	12.91...	TOC	..

Method ID	Sample Type	Vial	Timestamp	Message
TOC 0 - 20 ppmC	TOC Standard	1	2021/03/01 14:46	
TOC 0 - 20 ppmC	TOC Standard	2	2021/03/01 15:09	
TOC 0 - 20 ppmC	TOC Standard	3	2021/03/01 15:32	
TOC 0 - 20 ppmC	...TOC Standard	.. 4..	2021/03/01 15:55	..
TOC 0 - 20 ppmC	TOC Standard	5	2021/03/01 16:19	
TOC 0 - 20 ppmC	TOC Standard	6	2021/03/01 16:44	
TOC 0 - 20 ppmC	...TOC Standard	.. 7..	2021/03/01 17:11	..
TOC 0 - 20 ppmC	Sample	8	2021/03/01 17:36	
TOC 0 - 20 ppmC	Sample	9	2021/03/01 17:58	
TOC 0 - 20 ppmC	...Sample	.. 10..	2021/03/01 18:20	..

```
=====
Sample ID:   0.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   03011425
Cal. Curve:  TOC WATER 9060A            Timestamp: 2021/03/01 14:46
Operator ID: AM AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			10294	-0.859	-0.664	99
2			9525	-0.848	-0.651	99
3			9903	-0.852	-0.654	98
4			14018	-0.871	-0.672	101

```
-----
<<<Statistics>>>   Mean:    10935   Std Dev:    2079   RSD: 19.01
=====
```

```
=====
Sample ID:   0.5PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   03011425
Cal. Curve:  TOC WATER 9060A            Timestamp: 2021/03/01 15:09
Operator ID: AM AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			30892	-0.832	-0.637	106
2			28439	-0.813	-0.615	106
3			31486	-0.832	-0.632	106
4			30584	-0.819	-0.621	107

```
-----
<<<Statistics>>>   Mean:    30350   Std Dev:    1328   RSD: 4.38
=====
```

```
=====
Sample ID:   1.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   03011425
Cal. Curve:  TOC WATER 9060A            Timestamp: 2021/03/01 15:32
Operator ID: AM AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			48722	-0.789	-0.593	111
2			49536	-0.766	-0.568	111
3			53756	-0.803	-0.605	112
4			48247	-0.776	-0.576	109

```
-----
<<<Statistics>>>   Mean:    50065   Std Dev:    2517   RSD: 5.03
=====
```

```
=====
Sample ID:   2.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   03011425
Cal. Curve:  TOC WATER 9060A            Timestamp: 2021/03/01 15:55
Operator ID: AM AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1			87722	-0.813	-0.613	114
2			85972	-0.798	-0.599	114
3			88535	-0.801	-0.602	115
4			92615	-0.814	-0.617	115

```
-----
<<<Statistics>>>   Mean:    88711   Std Dev:    2814   RSD: 3.17
=====
```

```
=====
Sample ID:   5.0PPM                      Mode:      TOC
Method:      TOC 0 - 20 ppmC             Filename:   03011425
Cal. Curve:  TOC WATER 9060A            Timestamp: 2021/03/01 16:19
Operator ID: AM AP                      Sample Type: TOC Standard
=====
```

Rep #	ppm C	ug C	Raw Data	Beginning	Ending	Integration
-------	-------	------	----------	-----------	--------	-------------

			Baseline	Baseline	Time
1		204787	-0.783	-0.585	125
2		203365	-0.732	-0.534	124
3		201606	-0.708	-0.508	124
4		201438	-0.712	-0.514	121
<<<Statistics>>> Mean: 202799 Std Dev: 1586 RSD: 0.78					
=====					
Sample ID: 10.0PPM			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 16:44		
Operator ID: AM AP			Sample Type: TOC Standard		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1			444583	-0.834	-0.635 140
2			444269	-0.741	-0.544 156
3			434212	-0.749	-0.549 131
4			445526	-0.764	-0.567 133
<<<Statistics>>> Mean: 442148 Std Dev: 5317 RSD: 1.20					
=====					
Sample ID: 20.0PPM			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:11		
Operator ID: AM AP			Sample Type: TOC Standard		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1			871993	-0.802	-0.604 171
2			867580	-0.709	-0.511 171
3			861473	-0.666	-0.466 160
4			871494	-0.636	-0.440 160
<<<Statistics>>> Mean: 868135 Std Dev: 4860 RSD: 0.56					
=====					
Sample ID: ICV			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:36		
Operator ID: AM AP			Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1	10.6416	5.3208	458692	-0.807	-0.608 148
2	10.1032	5.0516	435485	-0.730	-0.531 140
3	10.5158	5.2579	453270	-0.672	-0.472 133
4	10.4286	5.2143	449513	-0.723	-0.523 145
<<<Statistics>>> Mean: 10.4223 Std Dev: 0.2300 RSD: 2.21					
=====					
Sample ID: ICB			Mode: TOC		
Method: TOC 0 - 20 ppmC			Filename: 03011425		
Cal. Curve: TOC WATER 9060A			Timestamp: 2021/03/01 17:58		
Operator ID: AM AP			Sample Type: Sample		
Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline Integration Time
1	0.2127	0.1064	9168	-0.833	-0.637 99
2	0.2987	0.1493	12875	-0.846	-0.647 100
3	0.2141	0.1071	9229	-0.812	-0.614 98
4	0.2711	0.1355	11684	-0.856	-0.657 100
<<<Statistics>>> Mean: 0.2492 Std Dev: 0.0428 RSD: 17.18					
=====					

Sample ID: IC-20PPM
Method: TOC 0 - 20 ppmC
Cal. Curve: TOC WATER 9060A
Operator ID: AM AP

Mode: TOC
Filename: 03011425
Timestamp: 2021/03/01 18:20
Sample Type: Sample

Rep #	ppm C	ug C	Raw Data	Beginning Baseline	Ending Baseline	Integration Time
1	0.2865	0.1433	12349	-0.902	-0.703	100
2	0.2592	0.1296	11171	-0.923	-0.727	100
3	0.2321	0.1161	10006	-0.935	-0.736	99
4	0.3140	0.1570	13534	-0.931	-0.731	101

<<<Statistics>>> Mean: 0.2729 Std Dev: 0.0352 RSD: 12.91

Cal. Curve ID: TOC WATER 9060A
 Created: 2021/03/01 17:11
 Calibration Factor (m): 8.621e+04
 Y Intercept (b): 4807
 r-squared: 0.99930

Standard ID	Y Raw Data	X Expected ug C	Measured ug C	Message	Date & Time
0.0PPM	10935	0.000	0.071	18	2021/03/01 14:46
0.5PPM	30350	0.250	0.296	5	2021/03/01 15:09
1.0PPM	50065	0.500	0.525	3	2021/03/01 15:32
2.0PPM	88711	1.000	0.973	8	2021/03/01 15:55
5.0PPM	202799	2.500	2.297	i	2021/03/01 16:19
10.0PPM	442148	5.000	5.073	0	2021/03/01 16:44
20.0PPM	868135	10.000	10.015		2021/03/01 17:11

03/02/2021
 AM