



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114083

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/21/2021	09:35
2	CAL2	1	Water	NA	NA	20.2	7.00	04/21/2021	09:36
3	CAL3	1	Water	NA	NA	20.2	10.02	04/21/2021	09:37
4	ICV	1	Water	NA	NA	20.3	7.01	04/21/2021	09:39
5	CCV1	1	Water	NA	NA	20.2	2.01	04/21/2021	09:40
6	M2114-01	1	Water	NA	NA	20.1	6.73	04/21/2021	09:42
7	M2114-02	1	Water	NA	NA	20.2	6.67	04/21/2021	09:44
8	M2114-03	1	Water	NA	NA	20.5	6.90	04/21/2021	09:46
9	M2114-04	1	Water	NA	NA	20.5	6.95	04/21/2021	09:48
10	M2114-07	1	Water	NA	NA	20.5	5.45	04/21/2021	09:49
11	M2115-01	1	Water	NA	NA	20.2	5.89	04/21/2021	09:50
12	M2115-01DUP	1	Water	NA	NA	20.3	5.90	04/21/2021	09:51
13	CCV2	1	Water	NA	NA	20.3	12.02	04/21/2021	10:00

HB 114083

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2114 WorkList ID : 148233 Department : Wet-Chemistry Date : 04-21-2021 09:18:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2021	Water	M2114-01	12 pH	Cool 4 deg C	WOOD06	O52	GMW-1	04/20/2021	9040C
04/20/2021	Water	M2114-02	12 pH	Cool 4 deg C	WOOD06	O52	GMW-10	04/20/2021	9040C
04/20/2021	Water	M2114-03	5 pH	Cool 4 deg C	WOOD06	O52	GMW-7	04/20/2021	9040C
04/20/2021	Water	M2114-04	11 pH	Cool 4 deg C	WOOD06	O52	GMW-X	04/20/2021	9040C
04/20/2021	Water	M2114-07	11 pH	Cool 4 deg C	WOOD06	O52	FB-02	04/20/2021	9040C
04/19/2021	Water	M2115-01	17 pH	Cool 4 deg C	STER05	P11	BOTTLE-1	04/19/2021	9040C

Date/Time 04/21/21 12:50 09:30
Raw Sample Received by: JO (wet chem) [Signature] 5x1
Raw Sample Relinquished by: [Signature] 5x1

Date/Time 04/21/21 11:00
Raw Sample Received by: [Signature] 5x1
Raw Sample Relinquished by: JO (wet chem)