



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114085

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/21/2021	09:35
2	CAL2	1	Water	NA	NA	20.2	7.00	04/21/2021	09:36
3	CAL3	1	Water	NA	NA	20.2	10.02	04/21/2021	09:37
4	ICV	1	Water	NA	NA	20.3	7.01	04/21/2021	09:39
5	CCV1	1	Water	NA	NA	20.2	2.01	04/21/2021	09:40
6	M2125-02	1	Water	NA	NA	20.6	11.77	04/21/2021	10:05
7	M2125-02DUP	1	Water	NA	NA	20.7	11.78	04/21/2021	10:06
8	CCV2	1	Water	NA	NA	20.3	12.02	04/21/2021	10:07

WORKLIST(Hardcopy Internal Chain)

183114085

WorkList Name : corrosivity m2125 WorkList ID : 148239 Department : Wet-Chemistry Date : 04-21-2021 11:30:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2021	Solid	M2125-01 - C	Corrosivity	Cool 4 deg C	ARCA05	P12	PB-2021-WS-000-07	04/20/2021	9045D

Date/Time 04/21/21 11:35
 Raw Sample Received by: JP (wet) (2pm)
 Raw Sample Relinquished by: JP (wet) (2pm)

Date/Time 04/21/21 12:35
 Raw Sample Received by: JP (wet) (2pm)
 Raw Sample Relinquished by: JP (wet) (2pm)