



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114086

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/21/2021	09:35
2	CAL2	1	Water	NA	NA	20.2	7.00	04/21/2021	09:36
3	CAL3	1	Water	NA	NA	20.2	10.02	04/21/2021	09:37
4	ICV	1	Water	NA	NA	20.3	7.01	04/21/2021	09:39
5	CCV1	1	Water	NA	NA	20.2	2.01	04/21/2021	09:40
6	M2125-01	1	Solid	20.03	20	20.7	11.83	04/21/2021	10:10
7	M2125-01DUP	1	Solid	20.04	20	20.8	11.84	04/21/2021	10:11
8	CCV2	1	Water	NA	NA	20.3	12.02	04/21/2021	10:15

WORKLIST(Hardcopy Internal Chain)

114086

WorkList Name : corrosivity m2115 WorkList ID : 148237 Department : Wet-Chemistry Date : 04-21-2021 11:24:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2021	Water	M2125-02	Corrosivity	Cool 4 deg C	ARCA05	P12	PB-2021-WWV-000-07	04/20/2021	9040C

Date/Time 04/21/21 11:35
 Raw Sample Received by: JH (wet Chem)
 Raw Sample Relinquished by: JH (wet Chem)

Date/Time 04/21/21 12:35
 Raw Sample Received by: JH (wet Chem)
 Raw Sample Relinquished by: JH (wet Chem)