



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114095

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	04/21/2021	12:20
2	M2101-04		1	15	100.00	>212.0	>212.0	04/21/2021	12:45
3	M2101-04DUP		1	15	100.00	>212.0	>212.0	04/21/2021	13:10
4	M2101-05		1	14	100.00	>212.0	>212.0	04/21/2021	13:35
5	M2101-06		1	15	100.00	>212.0	>212.0	04/21/2021	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114095

WorkList Name : fp42121 WorkList ID : 148231 Department : Wet-Chemistry Date : 04-21-2021 09:17:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/27/2021	Water	M2101-04	Flash Point	Cool 4 deg C	PSEG03	O32	30834	04/20/2021	1010A
04/27/2021	Water	M2101-05	Flash Point	Cool 4 deg C	PSEG03	O32	31044	04/20/2021	1010A
04/27/2021	Water	M2101-06	Flash Point	Cool 4 deg C	PSEG03	O32	31056	04/20/2021	1010A

Date/Time 04/21/21 12:00
 Raw Sample Received by: KS (wetcher)
 Raw Sample Relinquished by: JP (wetcher)

Date/Time 04/21/21 14:15
 Raw Sample Received by: JP (wetcher)
 Raw Sample Relinquished by: KS (wetcher)