


LB114104

7470A

INSTRUMENT ID: CV1

 μ Abs.:

36693



A= 0.0000e+000

B= 2.7216e-004

C= 2.0737e-002

Rho= 0.9998913

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.047	0.047	95	0.000	95					
0.2	0.200	0.227	0.027	757	0.0 %	757					14
2.5	2.500	2.441	-0.059	8892	0.0 %	8892					-2
5.0	5.000	4.915	-0.085	17982	0.0 %	17982					-2
7.5	7.500	7.564	0.064	27715	0.0 %	27715					1
10.0	10.000	10.007	0.007	36693	0.0 %	36693					0

LB114104 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Apr 2021 08:59:48

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	95	-	0.0000	7470A	PPB	22 Apr 2021 09:04:32	S	Std
0.2 - 1	SP.2	757	-	0.2000	7470A	PPB	22 Apr 2021 09:06:48	S	Std
2.5 - 1	SP.5	8892	-	2.5000	7470A	PPB	22 Apr 2021 09:09:05	S	Std
5.0 - 1	SP.5	17982	-	5.0000	7470A	PPB	22 Apr 2021 09:11:22	S	Std
7.5 - 1	SP.5	27715	-	7.5000	7470A	PPB	22 Apr 2021 09:18:57	S	Std
10.0 - 1	SP.5	36693	-	10.0000	7470A	PPB	22 Apr 2021 09:26:34	S	Std
ICV99 - 1	ICV99	14374	3.9328	-	7470A	PPB	22 Apr 2021 09:29:17	U	SMPL
ICB99 - 1	ICB99	-153	-0.0209	-	7470A	PPB	22 Apr 2021 09:31:32	U	SMPL
CCV68 - 1	CCV68	17205	4.7033	-	7470A	PPB	22 Apr 2021 09:33:47	U	SMPL
CCB68 - 1	CCB68	-147	-0.0193	-	7470A	PPB	22 Apr 2021 09:36:02	U	SMPL
CRA - 1	CRA	708	0.2134	-	7470A	PPB	22 Apr 2021 09:38:16	U	SMPL
HighStd - 1	HighStd	36517	9.9594	-	7470A	PPB	22 Apr 2021 09:40:31	U	SMPL
ChkStd - 1	ChkStd	26081	7.1191	-	7470A	PPB	22 Apr 2021 09:42:46	U	SMPL
PB135907BL - 1	PBW	-247	-0.0465	-	7470A	PPB	22 Apr 2021 09:45:03	U	SMPL
PB135907BS - 1	LCSW	15357	4.2004	-	7470A	PPB	22 Apr 2021 09:47:20	U	SMPL
M2102-02 - 1	RO-124037	-141	-0.0176	-	7470A	PPB	22 Apr 2021 09:49:36	U	SMPL
M2102-04 - 1	RB37708RT	23	0.0270	-	7470A	PPB	22 Apr 2021 09:51:53	U	SMPL
M2102-06 - 1	RB44882RT	41	0.0319	-	7470A	PPB	22 Apr 2021 09:54:09	U	SMPL
M2103-02 - 1	RO-124030	35	0.0303	-	7470A	PPB	22 Apr 2021 09:56:26	U	SMPL
M2106-02 - 1	SB-22-COMP	50	0.0343	-	7470A	PPB	22 Apr 2021 09:58:42	U	SMPL
CCV69 - 1	CCV69	18046	4.9322	-	7470A	PPB	22 Apr 2021 10:00:56	U	SMPL
CCB69 - 1	CCB69	-121	-0.0122	-	7470A	PPB	22 Apr 2021 10:03:12	U	SMPL
M2106-02DUP - 1	SB-22-COMP	12	0.0240	-	7470A	PPB	22 Apr 2021 10:05:27	U	SMPL
M2106-02MS - 1	SB-22-COMPMS	15473	4.2319	-	7470A	PPB	22 Apr 2021 10:07:42	U	SMPL
M2106-02MSD - 1	SB-22-COMPMSD	14718	4.0264	-	7470A	PPB	22 Apr 2021 10:12:28	U	SMPL
M2106-05 - 1	SB-23-COMP	203	0.0760	-	7470A	PPB	22 Apr 2021 10:14:43	U	SMPL
M2106-08 - 1	SB-24-COMP	4	0.0218	-	7470A	PPB	22 Apr 2021 10:17:00	U	SMPL
M2106-11 - 1	SB-25-COMP	23	0.0270	-	7470A	PPB	22 Apr 2021 10:19:16	U	SMPL
M2106-14 - 1	SB-26-COMP	33	0.0297	-	7470A	PPB	22 Apr 2021 10:21:32	U	SMPL
M2106-17 - 1	SB-27-COMP	45	0.0330	-	7470A	PPB	22 Apr 2021 10:23:49	U	SMPL
M2106-20 - 1	SB-28-COMP	55	0.0357	-	7470A	PPB	22 Apr 2021 10:26:06	U	SMPL
M2107-02 - 1	SB-29-COMP	58	0.0365	-	7470A	PPB	22 Apr 2021 10:28:21	U	SMPL
CCV70 - 1	CCV70	18475	5.0490	-	7470A	PPB	22 Apr 2021 10:30:36	U	SMPL
CCB70 - 1	CCB70	-130	-0.0146	-	7470A	PPB	22 Apr 2021 10:32:52	U	SMPL
M2107-05 - 1	SB-30-COMP	4	0.0218	-	7470A	PPB	22 Apr 2021 10:35:07	U	SMPL
M2107-09 - 1	SB-31-COMP	42	0.0322	-	7470A	PPB	22 Apr 2021 10:37:23	U	SMPL
M2107-12 - 1	SB-32-COMP	44	0.0327	-	7470A	PPB	22 Apr 2021 10:39:38	U	SMPL
M2107-15 - 1	SB-33-COMP	48	0.0338	-	7470A	PPB	22 Apr 2021 10:41:54	U	SMPL
M2107-18 - 1	SB-34-COMP	55	0.0357	-	7470A	PPB	22 Apr 2021 10:44:09	U	SMPL
M2107-21 - 1	SB-35-COMP	74	0.0409	-	7470A	PPB	22 Apr 2021 10:46:25	U	SMPL
PB135907TB - 1	PB135878TB	-11	0.0177	-	7470A	PPB	22 Apr 2021 10:48:42	U	SMPL
M2106-02LX5 - 1		17	0.0254	-	7470A	PPB	22 Apr 2021 10:50:58	U	SMPL
M2106-02A - 1		15411	4.2151	-	7470A	PPB	22 Apr 2021 11:03:06	U	SMPL
CCV71 - 1	CCV71	18834	5.1467	-	7470A	PPB	22 Apr 2021 11:05:21	U	SMPL
CCB71 - 1	CCB71	-171	-0.0258	-	7470A	PPB	22 Apr 2021 11:07:37	U	SMPL