


LB114105

7471B

INSTRUMENT ID: CV1

 μ Abs.:

53919



Linear

A= 0.0000e+000

B= 1.8541e-004

C= -3.4374e-002

Rho= 0.9999776

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	140	0.000	140					
0.2	0.200	0.195	-0.005	1238	0.0 %	1238					
2.5	2.500	2.487	-0.013	13599	0.0 %	13599					
5.0	5.000	5.030	0.030	27312	0.0 %	27312					
7.5	7.500	7.534	0.034	40818	0.0 %	40818					
10.0	10.000	9.963	-0.037	53919	0.0 %	53919					

100
-3
-1
0
0

Method: 7471B Operator: Admin Date of Analysis: 22 Apr 2021 11:37:33

Date of Analysis: 22 Apr 2021 11:37:33

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.00007471B	PPB	22 Apr 2021 11:41:34	S	Std
0.2 - 1	50.2	1238	-	0.20007471B	PPB	22 Apr 2021 11:43:50	S	Std
2.5 - 1	52.5	13599	-	2.50007471B	PPB	22 Apr 2021 11:46:08	S	Std
5.0 - 1	55	27312	-	5.00007471B	PPB	22 Apr 2021 11:48:24	S	Std
7.5 - 1	57.5	40818	-	7.50007471B	PPB	22 Apr 2021 11:50:41	S	Std
10.0 - 1	510	53919	-	10.00007471B	PPB	22 Apr 2021 11:55:32	S	Std
ICV01 - 1	ICV01	21380	3.9297	-7471B	PPB	22 Apr 2021 11:58:20	U	SMPL
ICB01 - 1	ICB01	55	-0.0242	-7471B	PPB	22 Apr 2021 12:00:35	U	SMPL
CCV72 - 1	CCV72	28199	5.1941	-7471B	PPB	22 Apr 2021 12:02:51	U	SMPL
CCB72 - 1	CCB72	21	-0.0305	-7471B	PPB	22 Apr 2021 12:05:06	U	SMPL
CRA - 1	CRA	1165	0.1816	-7471B	PPB	22 Apr 2021 12:07:21	U	SMPL
HighStd - 1	HighStd	53877	9.9551	-7471B	PPB	22 Apr 2021 12:09:36	U	SMPL
ChkStd - 1	ChkStd	37300	6.8815	-7471B	PPB	22 Apr 2021 12:11:51	U	SMPL
PB135908BL - 1	PBS	135	-0.0093	-7471B	PPB	22 Apr 2021 12:16:59	U	SMPL
PB135908BS - 1	LCSS	20627	3.7901	-7471B	PPB	22 Apr 2021 12:19:15	U	SMPL
M2099-01 - 1	TP-1	6981	1.2600	-7471B	PPB	22 Apr 2021 12:21:31	U	SMPL
M2099-03 - 1	TP-2	6104	1.0974	-7471B	PPB	22 Apr 2021 12:23:48	U	SMPL
M2099-05 - 1	TP-3	5672	1.0173	-7471B	PPB	22 Apr 2021 12:26:05	U	SMPL
M2099-07 - 1	TP-4	7203	1.3011	-7471B	PPB	22 Apr 2021 12:28:22	U	SMPL
M2099-09 - 1	TP-5	14199	2.5983	-7471B	PPB	22 Apr 2021 12:30:36	U	SMPL
CCV73 - 1	CCV73	28148	5.1846	-7471B	PPB	22 Apr 2021 12:32:52	U	SMPL
CCB73 - 1	CCB73	72	-0.0210	-7471B	PPB	22 Apr 2021 12:35:07	U	SMPL
M2103-01 - 1	RO-124030	5021	0.8966	-7471B	PPB	22 Apr 2021 12:37:23	U	SMPL
M2106-01 - 1	SB-22-0-2.0	2726	0.4711	-7471B	PPB	22 Apr 2021 12:39:38	U	SMPL
M2106-03 - 1	SB-22-14.5-15.0	9446	1.7170	-7471B	PPB	22 Apr 2021 12:41:53	U	SMPL
M2106-04 - 1	SB-23-0-2.0	6339	1.1410	-7471B	PPB	22 Apr 2021 12:44:09	U	SMPL
M2106-06 - 1	SB-23-14.5-15.0	561	0.0696	-7471B	PPB	22 Apr 2021 12:46:24	U	SMPL
M2106-07 - 1	SB-24-0-2.0	3747	0.6604	-7471B	PPB	22 Apr 2021 12:48:40	U	SMPL
M2106-09 - 1	SB-24-14.5-15.0	13137	2.4014	-7471B	PPB	22 Apr 2021 12:50:56	U	SMPL
M2106-10 - 1	SB-25-0-2.0	7134	1.2884	-7471B	PPB	22 Apr 2021 12:53:13	U	SMPL
M2106-12 - 1	SB-25-9.5-10.0	1233	0.1942	-7471B	PPB	22 Apr 2021 12:55:30	U	SMPL
M2106-13 - 1	SB-26-0-2.0	4986	0.8901	-7471B	PPB	22 Apr 2021 12:57:46	U	SMPL
CCV74 - 1	CCV74	28674	5.2821	-7471B	PPB	22 Apr 2021 13:00:01	U	SMPL
CCB74 - 1	CCB74	151	-0.0064	-7471B	PPB	22 Apr 2021 13:02:17	U	SMPL
M2106-15 - 1	SB-26-7.5-8.0	410	0.0416	-7471B	PPB	22 Apr 2021 13:04:32	U	SMPL
M2106-16 - 1	SB-27-0-2.0	6319	1.1372	-7471B	PPB	22 Apr 2021 13:06:48	U	SMPL
M2106-18 - 1	SB-27-9.5-10.0	74	-0.0207	-7471B	PPB	22 Apr 2021 13:09:03	U	SMPL
M2106-19 - 1	SB-28-0-2.0	7170	1.2950	-7471B	PPB	22 Apr 2021 13:11:19	U	SMPL
M2106-21 - 1	SB-28-9.5-10.0	656	0.0873	-7471B	PPB	22 Apr 2021 13:13:35	U	SMPL
M2106-21DUP - 1	SB-28-9.5-10.0DUP	476	0.0539	-7471B	PPB	22 Apr 2021 13:15:51	U	SMPL
M2106-21MS - 1	SB-28-9.5-10.0MS	20520	3.7703	-7471B	PPB	22 Apr 2021 13:18:08	U	SMPL
M2106-21MSD - 1	SB-28-9.5-10.0MSD	20158	3.7032	-7471B	PPB	22 Apr 2021 13:20:24	U	SMPL
PB135909BL - 1	PBS	-40	-0.0418	-7471B	PPB	22 Apr 2021 13:22:41	U	SMPL
PB135909BS - 1	LCSS	19578	3.5956	-7471B	PPB	22 Apr 2021 13:24:57	U	SMPL
CCV75 - 1	CCV75	28637	5.2753	-7471B	PPB	22 Apr 2021 13:27:12	U	SMPL
CCB75 - 1	CCB75	101	-0.0156	-7471B	PPB	22 Apr 2021 13:29:28	U	SMPL
M2102-01 - 1	RO-124037	1130	0.1751	-7471B	PPB	22 Apr 2021 13:31:44	U	SMPL
M2102-03 - 1	RB37708RT	7026	1.2683	-7471B	PPB	22 Apr 2021 13:33:59	U	SMPL
M2102-05 - 1	RB44882RT	3205	0.5599	-7471B	PPB	22 Apr 2021 13:36:14	U	SMPL
M2107-01 - 1	SB-29-0-2.0	5109	0.9129	-7471B	PPB	22 Apr 2021 13:45:42	U	SMPL
M2107-01DUP - 1	SB-29-0-2.0DUP	6024	1.0825	-7471B	PPB	22 Apr 2021 13:50:35	U	SMPL
M2107-01MS - 1	SB-29-0-2.0MS	26741	4.9237	-7471B	PPB	22 Apr 2021 13:52:51	U	SMPL
M2107-01MSD - 1	SB-29-0-2.0MSD	28219	5.1978	-7471B	PPB	22 Apr 2021 13:55:07	U	SMPL
M2107-03 - 1	SB-29-9.5-10.0	943	0.1405	-7471B	PPB	22 Apr 2021 13:57:24	U	SMPL
M2107-04 - 1	SB-30-0-2.0	14287	2.6146	-7471B	PPB	22 Apr 2021 13:59:41	U	SMPL
CCV76 - 1	CCV76	28434	5.2376	-7471B	PPB	22 Apr 2021 14:04:12	U	SMPL
CCB76 - 1	CCB76	141	-0.0082	-7471B	PPB	22 Apr 2021 14:06:27	U	SMPL
M2107-08 - 1	SB-31-0-2.0	4370	0.7759	-7471B	PPB	22 Apr 2021 14:10:58	U	SMPL
M2107-10 - 1	SB-31-10.5-11.0	6943	1.2529	-7471B	PPB	22 Apr 2021 14:13:13	U	SMPL
M2107-11 - 1	SB-32-0-2.0	4487	0.7976	-7471B	PPB	22 Apr 2021 14:15:28	U	SMPL
M2107-13 - 1	SB-32-10.5-11.0	14795	2.7088	-7471B	PPB	22 Apr 2021 14:17:44	U	SMPL
M2107-14 - 1	SB-33-0-2.0	4432	0.7874	-7471B	PPB	22 Apr 2021 14:20:00	U	SMPL
M2107-16 - 1	SB-33-10.5-11.0	52171	9.6387	-7471B	PPB	22 Apr 2021 14:22:16	U	SMPL
M2107-17 - 1	SB-34-0-2.0	4366	0.7751	-7471B	PPB	22 Apr 2021 14:24:32	U	SMPL
M2107-19 - 1	SB-34-9.5-10.0	13268	2.4257	-7471B	PPB	22 Apr 2021 14:26:50	U	SMPL
M2107-20 - 1	SB-35-0-2.0	6331	1.1395	-7471B	PPB	22 Apr 2021 14:29:06	U	SMPL
CCV77 - 1	CCV77	28546	5.2584	-7471B	PPB	22 Apr 2021 14:31:22	U	SMPL
CCB77 - 1	CCB77	154	-0.0058	-7471B	PPB	22 Apr 2021 14:33:38	U	SMPL
M2107-22 - 1	SB-35-10.5-11.0	18753	3.4427	-7471B	PPB	22 Apr 2021 14:35:54	U	SMPL
M2112-01 - 1	M-1	3484	0.6116	-7471B	PPB	22 Apr 2021 14:38:10	U	SMPL
M2112-02 - 1	M-2	585	0.0741	-7471B	PPB	22 Apr 2021 14:40:27	U	SMPL
PB135910BL - 1	PBS	7	-0.0331	-7471B	PPB	22 Apr 2021 14:42:42	U	SMPL
PB135910BS - 1	LCSS	22750	4.1837	-7471B	PPB	22 Apr 2021 14:44:58	U	SMPL
M2110-01 - 1	S-5	2251	0.3830	-7471B	PPB	22 Apr 2021 14:47:15	U	SMPL
M2110-02 - 1	S-6	737	0.1023	-7471B	PPB	22 Apr 2021 14:49:31	U	SMPL
M2110-03 - 1	S-7	1796	0.2986	-7471B	PPB	22 Apr 2021 14:51:47	U	SMPL
M2110-04 - 1	S-8	1914	0.3205	-7471B	PPB	22 Apr 2021 14:54:04	U	SMPL
M2111-01 - 1	S-1	1907	0.3192	-7471B	PPB	22 Apr 2021 14:56:20	U	SMPL
CCV78 - 1	CCV78	28629	5.2738	-7471B	PPB	22 Apr 2021 14:58:36	U	SMPL
CCB78 - 1	CCB78	152	-0.0062	-7471B	PPB	22 Apr 2021 15:00:52	U	SMPL
M2111-02 - 1	S-2	1712	0.2831	-7471B	PPB	22 Apr 2021 15:03:09	U	SMPL

LB114105 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Apr 2021 11:37:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
M2111-03 - 1	S-3	3425	0.6007		-7471B	PPB	22 Apr 2021 15:05:25	U	SMPL
M2111-04 - 1	S-4	3366	0.5897		-7471B	PPB	22 Apr 2021 15:07:41	U	SMPL
M2111-05 - 1	S-9	1885	0.3151		-7471B	PPB	22 Apr 2021 15:09:57	U	SMPL
M2111-06 - 1	S-10	6902	1.2453		-7471B	PPB	22 Apr 2021 15:12:13	U	SMPL
M2111-07 - 1	S-11	2225	0.3782		-7471B	PPB	22 Apr 2021 15:14:29	U	SMPL
M2111-07DUP - 1	S-11DUP	2107	0.3563		-7471B	PPB	22 Apr 2021 15:16:46	U	SMPL
M2111-07MS - 1	S-11MS	25539	4.7009		-7471B	PPB	22 Apr 2021 15:19:03	U	SMPL
M2111-07MSD - 1	S-11MSD	24322	4.4752		-7471B	PPB	22 Apr 2021 15:21:19	U	SMPL
M2117-01 - 1	PL-01-042121	4393	0.7801		-7471B	PPB	22 Apr 2021 15:23:36	U	SMPL
CCV79 - 1	CCV79	28268	5.2068		-7471B	PPB	22 Apr 2021 15:25:52	U	SMPL
CCB79 - 1	CCB79	-262	-0.0830		-7471B	PPB	22 Apr 2021 15:31:07	U	SMPL
M2118-01 - 1	339	880	0.1288		-7471B	PPB	22 Apr 2021 15:33:23	U	SMPL
M2118-03 - 1	20071	1147	0.1783		-7471B	PPB	22 Apr 2021 15:35:39	U	SMPL
M2119-01 - 1	12016	745	0.1038		-7471B	PPB	22 Apr 2021 15:37:55	U	SMPL
M2119-03 - 1	352	1837	0.3062		-7471B	PPB	22 Apr 2021 15:40:11	U	SMPL
M2128-01 - 1	COMP-1	2721	0.4701		-7471B	PPB	22 Apr 2021 15:42:27	U	SMPL
M2128-02 - 1	COMP-2	4164	0.7377		-7471B	PPB	22 Apr 2021 15:44:43	U	SMPL
M2107-06 - 1	SB-30-905-10.0	9366	1.7022		-7471B	PPB	22 Apr 2021 15:46:59	U	SMPL
M2107-07 - 1	SB-30-905-10.0UP	8009	1.4506		-7471B	PPB	22 Apr 2021 15:49:16	U	SMPL
CCV80 - 1	CCV80	28726	5.2918		-7471B	PPB	22 Apr 2021 15:51:33	U	SMPL
CCB80 - 1	CCB80	142	-0.0080		-7471B	PPB	22 Apr 2021 15:53:49	U	SMPL
M2106-21LX5 - 1		290	0.0194		-7471B	PPB	22 Apr 2021 15:56:06	U	SMPL
M2106-21A - 1		21953	4.0360		-7471B	PPB	22 Apr 2021 15:58:23	U	SMPL
M2107-01LX5 - 1		1589	0.2602		-7471B	PPB	22 Apr 2021 16:05:06	U	SMPL
M2107-01A - 1		26806	4.9358		-7471B	PPB	22 Apr 2021 16:07:21	U	SMPL
M2111-07LX5 - 1		551	0.0678		-7471B	PPB	22 Apr 2021 16:09:38	U	SMPL
M2111-07A - 1		24893	4.5811		-7471B	PPB	22 Apr 2021 16:11:55	U	SMPL
CCV81 - 1	CCV81	28143	5.1837		-7471B	PPB	22 Apr 2021 16:14:11	U	SMPL
CCB81 - 1	CCB81	120	-0.0121		-7471B	PPB	22 Apr 2021 16:16:27	U	SMPL