



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114106

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	81.2	04/22/2021	10:55
2	M2120-01		1	15	100.00	>212.0	>212.0	04/22/2021	11:20
3	M2120-01DUP		1	15	100.00	>212.0	>212.0	04/22/2021	11:45
4	M2125-02		1	12	100.00	>212.0	>212.0	04/22/2021	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP42221 WorkList ID : 148269 Department : Wet-Chemistry Date : 04-22-2021 10:27:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Water	M2120-01	C Flash Point	Cool 4 deg C	PSEG03	O12	FMI-121	04/21/2021	1010A
04/28/2021	Water	M2125-02	H Flash Point	Cool 4 deg C	ARCA05	P12	PB-2021-VWV-000-07	04/20/2021	1010A

Date/Time 04/22/21 10:30
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: _____

Date/Time 04/22/21 12:25
 Raw Sample Received by: _____
 Raw Sample Relinquished by: KS (Wetchem)