



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114111

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/22/2021	12:35
2	CAL2	1	Water	NA	NA	20.2	7.00	04/22/2021	12:36
3	CAL3	1	Water	NA	NA	20.2	10.03	04/22/2021	12:37
4	ICV	1	Water	NA	NA	20.3	6.99	04/22/2021	12:39
5	CCV1	1	Water	NA	NA	20.2	2.01	04/22/2021	12:40
6	M2131-02	1	Water	NA	NA	20.5	8.59	04/22/2021	12:42
7	M2131-02DUP	1	Water	NA	NA	20.6	8.60	04/22/2021	12:43
8	CCV2	1	Water	NA	NA	20.3	12.02	04/22/2021	12:45

2B 114111

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion m2131 WorkList ID : 148272 Department : Wet-Chemistry Date : 04-22-2021 12:29:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/21/2021	Water	M2131-02	F Corrosivity	Cool 4 deg C	JACO05	P12	IDW-AQ-FT1-042121	04/21/2021	9040C

g.s.a.

Date/Time 04/22/21 12:30
Raw Sample Received by: Jp (wet gem)
Raw Sample Relinquished by: Jp

Date/Time 04/22/21 12:30
Raw Sample Received by: Jp
Raw Sample Relinquished by: Jp