



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114117

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	04/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.4	10.03	04/22/2021	16:16
4	ICV	1	Water	NA	NA	20.4	7.01	04/22/2021	16:17
5	CCV1	1	Water	NA	NA	20.3	2.01	04/22/2021	16:19
6	M2140-01	1	Solid	20.02	20	20.3	5.90	04/22/2021	16:22
7	M2140-01DUP	1	Solid	20.03	20	20.4	5.91	04/22/2021	16:23
8	M2140-03	1	Solid	20.04	20	20.6	5.79	04/22/2021	16:25
9	M2141-01	1	Solid	20.04	20	20.6	5.56	04/22/2021	16:26
10	M2141-03	1	Solid	20.01	20	20.2	6.55	04/22/2021	16:27
11	M2141-05	1	Solid	20.03	20	20.5	5.22	04/22/2021	16:30
12	CCV2	1	Water	NA	NA	20.3	12.02	04/22/2021	16:32

WORKLIST(Hardcopy Internal Chain)

20114117

WorkList Name: ph m240 solids WorkList ID: 148281 Department: Wet-Chemistry Date: 04-22-2021 15:45:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/22/2021	Solid	M2140-01	pH	Cool 4 deg C	PSEG03	O13	203	04/22/2021	9045D
04/22/2021	Solid	M2140-03	pH	Cool 4 deg C	PSEG03	O13	205	04/22/2021	9045D
04/22/2021	Solid	M2141-01	pH	Cool 4 deg C	PSEG03	O13	202	04/22/2021	9045D
04/22/2021	Solid	M2141-03	pH	Cool 4 deg C	PSEG03	O13	223	04/22/2021	9045D
04/22/2021	Solid	M2141-05	pH	Cool 4 deg C	PSEG03	O13	935	04/22/2021	9045D

Date/Time 04/22/21 16:05
 Raw Sample Received by: JO Cent Chem
 Raw Sample Relinquished by: AM Cwif-chem

Date/Time 04/22/21 17:45
 Raw Sample Received by: AM Cwif-chem
 Raw Sample Relinquished by: JO Cent Chem