



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114118

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	04/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.4	10.03	04/22/2021	16:16
4	ICV	1	Water	NA	NA	20.4	7.01	04/22/2021	16:17
5	CCV1	1	Water	NA	NA	20.2	2.01	04/22/2021	16:19
6	M2139-01	1	Solid	20.02	20	20.1	5.57	04/22/2021	16:20
7	M2140-01	1	Solid	20.02	20	20.3	5.90	04/22/2021	16:22
8	M2140-01DUP	1	Solid	20.03	20	20.4	5.91	04/22/2021	16:23
9	M2140-03	1	Solid	20.04	20	20.6	5.79	04/22/2021	16:25
10	M2149-02	1	Solid	20.03	20	20.1	7.28	04/22/2021	16:35
11	M2149-05	1	Solid	20.02	20	20.3	7.78	04/22/2021	16:36
12	M2149-08	1	Solid	20.03	20	20.6	7.66	04/22/2021	16:39
13	M2149-11	1	Solid	20.01	20	20.6	7.60	04/22/2021	16:40
14	M2149-14	1	Solid	20.03	20	20.5	7.85	04/22/2021	16:42
15	CCV2	1	Water	NA	NA	20.2	12.02	04/22/2021	16:44

28114118

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m2140

WorkList ID : 148280

Department : Wet-Chemistry

Date : 04-22-2021 15:45:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/22/2021	Solid	M2139-01	Corrosivity	Cool 4 deg C	PSEG03	O13	OIL-DEBRIS-1-2	04/22/2021	9045D
04/22/2021	Solid	M2140-01	Corrosivity	Cool 4 deg C	PSEG03	O13	203	04/22/2021	9045D
04/22/2021	Solid	M2140-03	Corrosivity	Cool 4 deg C	PSEG03	O13	205	04/22/2021	9045D
04/21/2021	Solid	M2149-02	Corrosivity	Cool 4 deg C	LIRO01	P13	SB-36-COMP	04/21/2021	9045D
04/21/2021	Solid	M2149-05	Corrosivity	Cool 4 deg C	LIRO01	P13	SB-37-COMP	04/21/2021	9045D
04/21/2021	Solid	M2149-08	Corrosivity	Cool 4 deg C	LIRO01	P13	SB-38-COMP	04/21/2021	9045D
04/21/2021	Solid	M2149-11	Corrosivity	Cool 4 deg C	LIRO01	P13	SB-39-COMP	04/21/2021	9045D
04/21/2021	Solid	M2149-14	Corrosivity	Cool 4 deg C	LIRO01	P13	SB-40-COMP	04/21/2021	9045D

Date/Time 04/22/21 16:05
 Raw Sample Received by: JO Wet Chem
 Raw Sample Relinquished by: APL Wet Chem

Date/Time 04/22/21 17:45
 Raw Sample Received by: APL Wet Chem
 Raw Sample Relinquished by: JO Wet Chem