



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114119

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	04/22/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	04/22/2021	16:16
3	CAL3	1	Water	NA	NA	20.4	10.03	04/22/2021	16:16
4	ICV	1	Water	NA	NA	20.4	7.01	04/22/2021	16:17
5	CCV1	1	Water	NA	NA	20.4	2.01	04/22/2021	16:45
6	M2120-01	1	Water	NA	NA	20.2	6.83	04/22/2021	16:47
7	M2136-02	1	Water	NA	NA	20.1	7.17	04/22/2021	16:48
8	M2136-03	1	Water	NA	NA	20.3	7.42	04/22/2021	16:49
9	M2136-05	1	Water	NA	NA	20.6	6.35	04/22/2021	16:52
10	M2136-06	1	Water	NA	NA	20.6	7.34	04/22/2021	17:00
11	M2136-06DUP	1	Water	NA	NA	20.7	7.35	04/22/2021	17:02
12	CCV2	1	Water	NA	NA	20.4	12.02	04/22/2021	17:05

WORKLIST(Hardcopy Internal Chain)

28116119

WorkList Name : ph m2140

WorkList ID : 148279

Department : Wet-Chemistry

Date : 04-22-2021 15:44:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/21/2021	Water	M2120-01	pH	Cool 4 deg C	PSEG03	O12	FMI-121	04/21/2021	9040C
04/22/2021	Water	M2136-02	pH	Cool 4 deg C	PSEG03	O21	31075	04/22/2021	9040C
04/22/2021	Water	M2136-03	pH	Cool 4 deg C	PSEG03	O21	31076	04/22/2021	9040C
04/22/2021	Water	M2136-05	pH	Cool 4 deg C	PSEG03	O21	31032	04/22/2021	9040C
04/22/2021	Water	M2136-06	pH	Cool 4 deg C	PSEG03	O21	41765	04/22/2021	9040C

Date/Time 04/22/21 16:05
 Raw Sample Received by: JP (wet-chem)
 Raw Sample Relinquished by: JPM (wet-chem)

Date/Time 04/22/21 17:45
 Raw Sample Received by: JPM (wet-chem)
 Raw Sample Relinquished by: JP (wet-chem)