


LB114128

7470A

INSTRUMENT ID: CVL

μ Abs.:

38414



Linear

A= 0.0000e+000

B= 2.6143e-004

C= 9.1548e-003

Rho= 0.9998239

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.086	0.086	293	0.000	293					-
0.2	0.200	0.223	0.023	818	0.0 %	818					12
2.5	2.500	2.424	-0.076	9236	0.0 %	9236					-3
5.0	5.000	4.891	-0.109	18675	0.0 %	18675					-2
7.5	7.500	7.524	0.024	28747	0.0 %	28747					0
10.0	10.000	10.052	0.052	38414	0.0 %	38414					1

LB114128 IN STRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Apr 2021 11:12:08					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	293	-	0.0000	7470A	PPB	23 Apr 2021 11:23:34	S	Std
0.2 - 1	50.2	818	-	0.2000	7470A	PPB	23 Apr 2021 11:25:50	S	Std
2.5 - 1	52.5	9236	-	2.5000	7470A	PPB	23 Apr 2021 11:28:08	S	Std
5.0 - 1	55	18675	-	5.0000	7470A	PPB	23 Apr 2021 11:30:25	S	Std
7.5 - 1	57.5	28747	-	7.5000	7470A	PPB	23 Apr 2021 11:32:41	S	Std
10.0 - 1	510	38414	-	10.0000	7470A	PPB	23 Apr 2021 11:34:56	S	Std
ICV02 - 1	ICV02	15855	4.1541	-	7470A	PPB	23 Apr 2021 11:37:42	U	SMPL
ICB02 - 1	ICB02	163	0.0518	-	7470A	PPB	23 Apr 2021 11:42:35	U	SMPL
CCV82 - 1	CCV82	20233	5.2987	-	7470A	PPB	23 Apr 2021 11:47:19	U	SMPL
CCB82 - 1	CCB82	15	0.0131	-	7470A	PPB	23 Apr 2021 11:49:33	U	SMPL
CRA - 1	CRA	790	0.2157	-	7470A	PPB	23 Apr 2021 11:51:48	U	SMPL
HighStd - 1	HighStd	39140	10.2415	-	7470A	PPB	23 Apr 2021 11:54:03	U	SMPL
ChkStd - 1	ChkStd	26397	6.9101	-	7470A	PPB	23 Apr 2021 11:56:18	U	SMPL
PB135962BL - 1	PBW	-186	-0.0395	-	7470A	PPB	23 Apr 2021 11:58:34	U	SMPL
PB135962BS - 1	LCSW	12949	3.3944	-	7470A	PPB	23 Apr 2021 12:00:52	U	SMPL
M2125-01 - 1	PB-2021-WS-000-07	-38	-0.0008	-	7470A	PPB	23 Apr 2021 12:03:08	U	SMPL
M2125-01DUP - 1	PB-2021-WS-000-07DUP	93	0.0335	-	7470A	PPB	23 Apr 2021 12:05:24	U	SMPL
M2125-01MS - 1	PB-2021-WS-000-07MS	14033	3.6778	-	7470A	PPB	23 Apr 2021 12:07:40	U	SMPL
M2125-01MSD - 1	PB-2021-WS-000-07MSD	14220	3.7267	-	7470A	PPB	23 Apr 2021 12:09:57	U	SMPL
PB135963BL - 1	PBW	9	0.0115	-	7470A	PPB	23 Apr 2021 12:12:12	U	SMPL
CCV83 - 1	CCV83	19263	5.0451	-	7470A	PPB	23 Apr 2021 12:14:27	U	SMPL
CCB83 - 1	CCB83	126	0.0421	-	7470A	PPB	23 Apr 2021 12:19:17	U	SMPL
PB135963BS - 1	LCSW	13464	3.5290	-	7470A	PPB	23 Apr 2021 12:21:32	U	SMPL
M2101-04 - 1	30834	2161	0.5741	-	7470A	PPB	23 Apr 2021 12:23:47	U	SMPL
M2114-01 - 1	GMW-1	16800	4.4012	-	7470A	PPB	23 Apr 2021 12:26:02	U	SMPL
M2114-02 - 1	GMW-10	4085	1.0771	-	7470A	PPB	23 Apr 2021 12:28:19	U	SMPL
M2114-03 - 1	GMW-7	2155	0.5725	-	7470A	PPB	23 Apr 2021 12:30:37	U	SMPL
M2114-04 - 1	GMW-X	3824	1.0089	-	7470A	PPB	23 Apr 2021 12:32:53	U	SMPL
M2114-04DUP - 1	GMW-XDUP	3717	0.9809	-	7470A	PPB	23 Apr 2021 12:35:09	U	SMPL
M2114-05 - 1	M2114-04MS	17600	4.6103	-	7470A	PPB	23 Apr 2021 12:37:25	U	SMPL
M2114-06 - 1	M2114-04MSD	17333	4.5405	-	7470A	PPB	23 Apr 2021 12:39:42	U	SMPL
M2114-07 - 1	FB-02	18	0.0139	-	7470A	PPB	23 Apr 2021 12:45:23	U	SMPL
CCV84 - 1	CCV84	18664	4.8885	-	7470A	PPB	23 Apr 2021 12:47:37	U	SMPL
CCB84 - 1	CCB84	-47	-0.0031	-	7470A	PPB	23 Apr 2021 12:49:53	U	SMPL
M2114-09 - 1	GMW-1	184	0.0573	-	7470A	PPB	23 Apr 2021 12:52:08	U	SMPL
M2114-10 - 1	GMW-10	116	0.0395	-	7470A	PPB	23 Apr 2021 12:54:23	U	SMPL
M2114-11 - 1	GMW-7	109	0.0377	-	7470A	PPB	23 Apr 2021 12:56:39	U	SMPL
M2114-12 - 1	GMW-X	184	0.0573	-	7470A	PPB	23 Apr 2021 12:58:55	U	SMPL
M2114-13 - 1	M2114-12MS	12544	3.2885	-	7470A	PPB	23 Apr 2021 13:01:11	U	SMPL
M2114-14 - 1	M2114-12MSD	12481	3.2721	-	7470A	PPB	23 Apr 2021 13:03:29	U	SMPL
M2114-15 - 1	FB-02	58	0.0243	-	7470A	PPB	23 Apr 2021 13:05:45	U	SMPL
M2120-01 - 1	FMI-121	482	0.1352	-	7470A	PPB	23 Apr 2021 13:08:02	U	SMPL
M2125-02 - 1	PB-2021-WW-000-07	1419	0.3801	-	7470A	PPB	23 Apr 2021 13:10:19	U	SMPL
M2131-02 - 1	IDW-AQ-FT1-042121	138	0.0452	-	7470A	PPB	23 Apr 2021 13:12:34	U	SMPL
CCV85 - 1	CCV85	18787	4.9206	-	7470A	PPB	23 Apr 2021 13:14:50	U	SMPL
CCB85 - 1	CCB85	-82	-0.0123	-	7470A	PPB	23 Apr 2021 13:17:05	U	SMPL
PB135962TB - 1	PB135902TB	111	0.0382	-	7470A	PPB	23 Apr 2021 13:19:21	U	SMPL
M2125-01LX5 - 1		24	0.0154	-	7470A	PPB	23 Apr 2021 13:21:36	U	SMPL
M2125-01A - 1		15141	3.9675	-	7470A	PPB	23 Apr 2021 13:23:52	U	SMPL
M2114-04LX5 - 1		537	0.1495	-	7470A	PPB	23 Apr 2021 13:26:06	U	SMPL
M2114-04A - 1		18602	4.8723	-	7470A	PPB	23 Apr 2021 13:28:22	U	SMPL
CCV86 - 1	CCV86	18905	4.9515	-	7470A	PPB	23 Apr 2021 13:30:38	U	SMPL
CCB86 - 1	CCB86	-101	-0.0172	-	7470A	PPB	23 Apr 2021 13:32:56	U	SMPL