


LB 114132

SFAM01.1

INSTRUMENT ID: CV1

μ Abs.:

43691



Linear

A= 0.0000e+000

B= 2.3037e-004

C= 1.1478e-002

Rho= 0.9999040

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.00	0.000	0.036	0.036	108	0.000	108					=
0.05	0.050					204					16
0.20	0.200	0.231	0.031	954	0.0 %	954					-2
2.50	2.500	2.459	-0.041	10625	0.0 %	10625					-1
5.00	5.000	4.964	-0.036	21499	0.0 %	21499					-1
7.50	7.500	7.432	-0.068	32213	0.0 %	32213					-1
10.0	10.000	10.077	0.077	43691	0.0 %	43691					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	108	-		0 SFAM01.1	PPB	4/23/2021 14:14	Std
	0.2 S01	954	-		0.2 SFAM01.1	PPB	4/23/2021 14:17	Std
	2.5 S02	10625	-		2.5 SFAM01.1	PPB	4/23/2021 14:19	Std
	5 S03	21499	-		5 SFAM01.1	PPB	4/23/2021 14:21	Std
	7.5 S04	32213	-		7.5 SFAM01.1	PPB	4/23/2021 14:23	Std
	10 S05	43691	-		10 SFAM01.1	PPB	4/23/2021 14:31	Std
ICV020	ICV020	15763	3.6428 -		SFAM01.1	PPB	4/23/2021 14:34	SMPL
ICB020	ICB020	-94	-0.0102 -		SFAM01.1	PPB	4/23/2021 14:36	SMPL
CCV086	CCV086	21420	4.946 -		SFAM01.1	PPB	4/23/2021 14:38	SMPL
CCB086	CCB086	-107	-0.0132 -		SFAM01.1	PPB	4/23/2021 14:40	SMPL
PB135964BL	PBW964	7	0.0131 -		SFAM01.1	PPB	4/23/2021 14:43	SMPL
M2058-01	MYAQY3	18	0.0156 -		SFAM01.1	PPB	4/23/2021 14:45	SMPL
M2058-02	MYAQZ1	11	0.014 -		SFAM01.1	PPB	4/23/2021 14:47	SMPL
M2058-03	MYAQZ3	208	0.0594 -		SFAM01.1	PPB	4/23/2021 14:49	SMPL
M2058-04	MYAR13	-5	0.0103 -		SFAM01.1	PPB	4/23/2021 14:52	SMPL
M2058-05	MYAR13D	44	0.0216 -		SFAM01.1	PPB	4/23/2021 14:54	SMPL
M2058-06	MYAR13S	4816	1.1209 -		SFAM01.1	PPB	4/23/2021 14:56	SMPL
M2058-07	MYAR21	8	0.0133 -		SFAM01.1	PPB	4/23/2021 14:58	SMPL
M2058-08	MYAR09	108	0.0364 -		SFAM01.1	PPB	4/23/2021 15:01	SMPL
M2058-09	MYAR11	113	0.0375 -		SFAM01.1	PPB	4/23/2021 15:03	SMPL
M2058-10	MYAR27	59	0.0251 -		SFAM01.1	PPB	4/23/2021 15:05	SMPL
M2058-11	MYAR65	22	0.0165 -		SFAM01.1	PPB	4/23/2021 15:07	SMPL
M2058-12	MYAR17	52	0.0235 -		SFAM01.1	PPB	4/23/2021 15:10	SMPL
M2058-13	MYAR19	31	0.0186 -		SFAM01.1	PPB	4/23/2021 15:12	SMPL
M2058-14	MYAR23	54	0.0239 -		SFAM01.1	PPB	4/23/2021 15:14	SMPL
M2058-15	MYAR25	24	0.017 -		SFAM01.1	PPB	4/23/2021 15:17	SMPL
M2058-16	MYAR55	93	0.0329 -		SFAM01.1	PPB	4/23/2021 15:19	SMPL
M2058-17	MYAR67	30	0.0184 -		SFAM01.1	PPB	4/23/2021 15:21	SMPL
M2058-18	MYAR69	51	0.0232 -		SFAM01.1	PPB	4/23/2021 15:23	SMPL
M2058-19	MYAR71	51	0.0232 -		SFAM01.1	PPB	4/23/2021 15:26	SMPL
CCV087	CCV087	20567	4.7495 -		SFAM01.1	PPB	4/23/2021 15:28	SMPL
CCB087	CCB087	-82	-0.0074 -		SFAM01.1	PPB	4/23/2021 15:30	SMPL
M2058-20	MYAR29	45	0.0218 -		SFAM01.1	PPB	4/23/2021 15:32	SMPL
M2058-21	MYAR37	0	0.0115 -		SFAM01.1	PPB	4/23/2021 15:35	SMPL
M2058-22	MYAR43	27	0.0177 -		SFAM01.1	PPB	4/23/2021 15:37	SMPL
PB135965BL	PBW965	27	0.0177 -		SFAM01.1	PPB	4/23/2021 15:39	SMPL
M2059-01	MYAQY4	46	0.0221 -		SFAM01.1	PPB	4/23/2021 15:41	SMPL
M2059-02	MYAQZ2	71	0.0278 -		SFAM01.1	PPB	4/23/2021 15:44	SMPL
M2059-03	MYAQZ4	232	0.0649 -		SFAM01.1	PPB	4/23/2021 15:46	SMPL
M2059-04	MYAR22	-29	0.0048 -		SFAM01.1	PPB	4/23/2021 15:48	SMPL
M2059-05	MYAR10	96	0.0336 -		SFAM01.1	PPB	4/23/2021 15:50	SMPL
M2059-06	MYAR12	130	0.0414 -		SFAM01.1	PPB	4/23/2021 15:53	SMPL
M2059-07	MYAR14	55	0.0241 -		SFAM01.1	PPB	4/23/2021 15:55	SMPL
M2059-08	MYAR14D	40	0.0207 -		SFAM01.1	PPB	4/23/2021 15:57	SMPL
M2059-09	MYAR14S	4272	0.9956 -		SFAM01.1	PPB	4/23/2021 16:02	SMPL
M2059-10	MYAR28	115	0.038 -		SFAM01.1	PPB	4/23/2021 16:04	SMPL

LB114132

INSTRUMENT ID : CV1

M2059-11	MYAR66	77	0.0292 -	SFAM01.1	PPB	4/23/2021 16:07 SMPL
M2059-12	MYAR18	49	0.0228 -	SFAM01.1	PPB	4/23/2021 16:09 SMPL
M2059-13	MYAR20	53	0.0237 -	SFAM01.1	PPB	4/23/2021 16:11 SMPL
M2059-14	MYAR24	36	0.0198 -	SFAM01.1	PPB	4/23/2021 16:13 SMPL
M2059-15	MYAR26	92	0.0327 -	SFAM01.1	PPB	4/23/2021 16:16 SMPL
M2059-16	MYAR56	621	0.1545 -	SFAM01.1	PPB	4/23/2021 16:18 SMPL
CCV088	CCV088	20216	4.6687 -	SFAM01.1	PPB	4/23/2021 16:20 SMPL
CCB088	CCB088	-79	-0.0067 -	SFAM01.1	PPB	4/23/2021 16:22 SMPL
M2059-17	MYAR68	117	0.0384 -	SFAM01.1	PPB	4/23/2021 16:25 SMPL
M2059-18	MYAR70	105	0.0357 -	SFAM01.1	PPB	4/23/2021 16:27 SMPL
M2059-19	MYAR72	46	0.0221 -	SFAM01.1	PPB	4/23/2021 16:29 SMPL
M2059-20	MYAR30	113	0.0375 -	SFAM01.1	PPB	4/23/2021 16:32 SMPL
M2059-21	MYAR38	58	0.0248 -	SFAM01.1	PPB	4/23/2021 16:34 SMPL
M2059-22	MYAR44	84	0.0308 -	SFAM01.1	PPB	4/23/2021 16:36 SMPL
CCV089	CCV089	20262	4.6792 -	SFAM01.1	PPB	4/23/2021 16:38 SMPL
CCB089	CCB089	-76	-0.006 -	SFAM01.1	PPB	4/23/2021 16:41 SMPL