



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114140

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/23/2021	13:35
2	M2131-02		1	15	100.00	>212.0	>212.0	04/23/2021	14:05
3	M2131-02DUP		1	15	100.00	>212.0	>212.0	04/23/2021	14:32
4	M2136-04		1	14	100.00	>212.0	>212.0	04/23/2021	14:58
5	M2136-05		1	12	100.00	>212.0	>212.0	04/23/2021	15:25
6	M2136-06		1	13	100.00	>212.0	>212.0	04/23/2021	15:50
7	M2150-01		1	15	100.00	>212.0	>212.0	04/23/2021	16:20
8	M2150-04		1	14	100.00	>212.0	>212.0	04/23/2021	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB114140

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp42321 WorkList ID : 148318 Department : Wet-Chemistry Date : 04-23-2021 09:04:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/04/2021	Water	M2131-02	Flash Point	Cool 4 deg C	JACO05	P12	IDW-AQ-FT1-042121	04/21/2021	1010A
04/29/2021	Water	M2136-04	Flash Point	Cool 4 deg C	PSEG03	O21	31075-76	04/22/2021	1010A
04/29/2021	Water	M2136-05	Flash Point	Cool 4 deg C	PSEG03	O21	31032	04/22/2021	1010A
04/29/2021	Water	M2136-06	Flash Point	Cool 4 deg C	PSEG03	O21	41765	04/22/2021	1010A
04/29/2021	Water	M2150-01	Flash Point	Cool 4 deg C	LIRO01	P12	TWP-20	04/22/2021	1010A
04/29/2021	Water	M2150-04	Flash Point	Cool 4 deg C	LIRO01	P12	TWP-40	04/22/2021	1010A

Date/Time 04/23/21 13:10
Raw Sample Received by: KS (Weber)
Raw Sample Relinquished by: CP (SMI)

Date/Time 04/23/21 16:50
Raw Sample Received by: CP (SMI)
Raw Sample Relinquished by: KS (Weber)