



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114142

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/23/2021	17:36
2	CAL2	1	Water	NA	NA	20.3	7.00	04/23/2021	17:37
3	CAL3	1	Water	NA	NA	20.4	10.03	04/23/2021	17:38
4	ICV	1	Water	NA	NA	20.4	7.00	04/23/2021	17:39
5	CCV1	1	Water	NA	NA	20.4	2.01	04/23/2021	17:40
6	M2170-08	1	Solid	20.03	20	20.3	8.13	04/23/2021	17:51
7	M2170-08DUP	1	Solid	20.04	20	20.3	8.14	04/23/2021	17:52
8	M2170-10	1	Solid	20.02	20	20.3	9.76	04/23/2021	18:00
9	M2170-12	1	Solid	20.04	20	20.1	8.82	04/23/2021	18:01
10	M2170-14	1	Solid	20.02	20	20.3	8.69	04/23/2021	18:02
11	M2170-16	1	Solid	20.04	20	20.4	8.73	04/23/2021	18:05
12	M2170-18	1	Solid	20.01	20	20.6	9.84	04/23/2021	18:10
13	CCV2	1	Water	NA	NA	20.3	12.02	04/23/2021	18:11

WORKLIST(Hardcopy Internal Chain)

11/1/21

WorkList Name : corrosivity m2170

WorkList ID : 148334

Department : Wet-Chemistry

Date : 04-23-2021 14:20:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/23/2021	Solid	M2170-08	Corrosivity	Cool 4 deg C	JPCL01	O31	DS-1	04/23/2021	9045D
04/23/2021	Solid	M2170-10	Corrosivity	Cool 4 deg C	JPCL01	O31	DN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-12	Corrosivity	Cool 4 deg C	JPCL01	O31	CS-1	04/23/2021	9045D
04/23/2021	Solid	M2170-14	Corrosivity	Cool 4 deg C	JPCL01	O31	CN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-16	Corrosivity	Cool 4 deg C	JPCL01	O31	BN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-18	Corrosivity	Cool 4 deg C	JPCL01	O31	AN-1	04/23/2021	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/23/21 14:30
JP (wet Chem)
JP (54)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/23/21 14:30
JP (54)
JP (wet Chem)