



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114143

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/23/2021	17:36
2	CAL2	1	Water	NA	NA	20.3	7.00	04/23/2021	17:37
3	CAL3	1	Water	NA	NA	20.4	10.03	04/23/2021	17:38
4	ICV	1	Water	NA	NA	20.4	7.00	04/23/2021	17:39
5	CCV1	1	Water	NA	NA	20.4	2.01	04/23/2021	17:40
6	M2154-02	1	Water	NA	NA	20.4	9.49	04/23/2021	18:20
7	M2154-02DUP	1	Water	NA	NA	20.5	9.50	04/23/2021	18:21
8	M2156-01	1	Water	NA	NA	20.6	5.44	04/23/2021	18:23
9	CCV2	1	Water	NA	NA	20.3	12.01	04/23/2021	18:25

WORKLIST(Hardcopy Internal Chain)

114143

WorkList Name : PH M2171 S

WorkList ID : 148336

Department : Wet-Chemistry

Date : 04-23-2021 16:11:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/23/2021	Water	M2154-02	pH	Cool 4 deg C	PSEG03	O23	33096	04/23/2021	9040C
04/23/2021	Water	M2156-01	pH	Cool 4 deg C	PSEG03	O31	33066	04/23/2021	9040C

Date/Time 04/23/21 16:20
 Raw Sample Received by: JO [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/23/21 17:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]