


LB114146

7471B

INSTRUMENT ID: CV1

μ Abs.:

48828



Linear

A= 0.0000e+000

B= 2.0468e-004

C= -1.7867e-002

Rho= 0.9998922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.001	0.001	93	0.000	93					900
0.2	0.200	0.207	0.007	1100	0.0 %	1100					1
2.5	2.500	2.426	-0.074	11941	0.0 %	11941					3
5.0	5.000	5.106	0.106	25034	0.0 %	25034					3
7.5	7.500	7.483	-0.017	36646	0.0 %	36646					2
10.0	10.000	9.976	-0.024	48828	0.0 %	48828					0

900
1
3
3
2
0
0

LB114146 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Apr 2021 09:46:25

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7471B	PPB	26 Apr 2021 09:51:17	S	Std
0.2 - 1	50.2	1100	-	0.2000	7471B	PPB	26 Apr 2021 09:53:34	S	Std
2.5 - 1	52.5	11941	-	2.5000	7471B	PPB	26 Apr 2021 09:55:52	S	Std
5.0 - 1	55	25034	-	5.0000	7471B	PPB	26 Apr 2021 09:58:09	S	Std
7.5 - 1	57.5	36646	-	7.5000	7471B	PPB	26 Apr 2021 10:00:25	S	Std
10.0 - 1	510	48828	-	10.0000	7471B	PPB	26 Apr 2021 10:02:42	S	Std
ICV03 - 1	ICV03	18728	3.8154	-	7471B	PPB	26 Apr 2021 10:10:23	U	SMPL
ICB03 - 1	ICB03	-97	-0.0377	-	7471B	PPB	26 Apr 2021 10:12:37	U	SMPL
CCV90 - 1	CCV90	25375	5.1759	-	7471B	PPB	26 Apr 2021 10:14:54	U	SMPL
CCB90 - 1	CCB90	-95	-0.0373	-	7471B	PPB	26 Apr 2021 10:17:09	U	SMPL
CRA - 1	CRA	1096	0.2065	-	7471B	PPB	26 Apr 2021 10:19:25	U	SMPL
HighStd - 1	HighStd	48316	9.8715	-	7471B	PPB	26 Apr 2021 10:21:40	U	SMPL
ChkStd - 1	ChkStd	34615	7.0672	-	7471B	PPB	26 Apr 2021 10:26:32	U	SMPL
PB135966BL - 1	PBS	-129	-0.0443	-	7471B	PPB	26 Apr 2021 10:28:47	U	SMPL
PB135966BS - 1	LCSS	18776	3.8252	-	7471B	PPB	26 Apr 2021 10:31:05	U	SMPL
M2132-01 - 1	TP-1-COMP	68113	13.9236	-	7471B	PPB	26 Apr 2021 10:33:21	U	SMPL
M2132-01DUP - 1	TP-1-COMPDUP	67600	13.8186	-	7471B	PPB	26 Apr 2021 10:35:38	U	SMPL
M2132-01MS - 1	TP-1-COMPMS	81058	16.5732	-	7471B	PPB	26 Apr 2021 10:41:17	U	SMPL
M2132-01MSD - 1	TP-1-COMPMSD	86743	17.7368	-	7471B	PPB	26 Apr 2021 10:43:34	U	SMPL
M2132-06 - 1	TP-2-COMP	533	0.0912	-	7471B	PPB	26 Apr 2021 10:45:51	U	SMPL
CCV91 - 1	CCV91	25253	5.1509	-	7471B	PPB	26 Apr 2021 10:48:09	U	SMPL
CCB91 - 1	CCB91	8	-0.0162	-	7471B	PPB	26 Apr 2021 10:50:25	U	SMPL
M2136-01 - 1	41763	7254	1.4669	-	7471B	PPB	26 Apr 2021 10:52:41	U	SMPL
M2142-01 - 1	EO-04-042221	3552	0.7092	-	7471B	PPB	26 Apr 2021 10:54:56	U	SMPL
M2143-01 - 1	HD-01-042221	2701	0.5350	-	7471B	PPB	26 Apr 2021 10:57:11	U	SMPL
M2145-01 - 1	SU-02-042221	273	0.0380	-	7471B	PPB	26 Apr 2021 10:59:27	U	SMPL
M2146-01 - 1	TR-04-042221	1251	0.2382	-	7471B	PPB	26 Apr 2021 11:01:42	U	SMPL
M2148-01 - 1	TP-18-17-COMP	1661	0.3221	-	7471B	PPB	26 Apr 2021 11:03:58	U	SMPL
M2152-01 - 1	TP-3-COMP	2882	0.5720	-	7471B	PPB	26 Apr 2021 11:06:14	U	SMPL
M2152-06 - 1	TP-4-COMP	32461	6.6263	-	7471B	PPB	26 Apr 2021 11:08:31	U	SMPL
M2163-01 - 1	TT-059-RW8-IDWSO-20210422-1	127	0.0081	-	7471B	PPB	26 Apr 2021 11:10:49	U	SMPL
PB135967BL - 1	PBS	63	-0.0050	-	7471B	PPB	26 Apr 2021 11:13:05	U	SMPL
CCV92 - 1	CCV92	25649	5.2320	-	7471B	PPB	26 Apr 2021 11:15:21	U	SMPL
CCB92 - 1	CCB92	26	-0.0125	-	7471B	PPB	26 Apr 2021 11:17:36	U	SMPL
PB135967BS - 1	LCSS	17729	3.6109	-	7471B	PPB	26 Apr 2021 11:19:52	U	SMPL
M2149-01 - 1	SB-36-0-2.0	7169	1.4495	-	7471B	PPB	26 Apr 2021 11:22:08	U	SMPL
M2149-01DUP - 1	SB-36-0-2.0DUP	7877	1.5944	-	7471B	PPB	26 Apr 2021 11:24:23	U	SMPL
M2149-01MS - 1	SB-36-0-2.0MS	25471	5.1956	-	7471B	PPB	26 Apr 2021 11:26:39	U	SMPL
M2149-01MSD - 1	SB-36-0-2.0MSD	26397	5.3851	-	7471B	PPB	26 Apr 2021 11:28:55	U	SMPL
M2149-03 - 1	SB-36-10.5-11.0	17752	3.6156	-	7471B	PPB	26 Apr 2021 11:31:12	U	SMPL
M2149-04 - 1	SB-37-0-2.0	1858	0.3624	-	7471B	PPB	26 Apr 2021 11:33:30	U	SMPL
M2149-06 - 1	SB-37-9.5-10.0	1819	0.3544	-	7471B	PPB	26 Apr 2021 11:35:47	U	SMPL
M2149-07 - 1	SB-38-0-2.0	43424	8.8702	-	7471B	PPB	26 Apr 2021 11:38:04	U	SMPL
M2149-09 - 1	SB-38-9.5-10.0	8582	1.7387	-	7471B	PPB	26 Apr 2021 11:40:19	U	SMPL
CCV93 - 1	CCV93	25248	5.1499	-	7471B	PPB	26 Apr 2021 11:42:36	U	SMPL
CCB93 - 1	CCB93	-1	-0.0181	-	7471B	PPB	26 Apr 2021 11:44:52	U	SMPL
M2149-10 - 1	SB-39-0-2.0	27552	5.6215	-	7471B	PPB	26 Apr 2021 11:47:08	U	SMPL
M2149-12 - 1	SB-39-9.5-10.0	294	0.0423	-	7471B	PPB	26 Apr 2021 11:49:24	U	SMPL
M2149-13 - 1	SB-40-0-2.0	6834	1.3809	-	7471B	PPB	26 Apr 2021 11:51:41	U	SMPL
M2149-15 - 1	SB-40-9.5-10.0	1001	0.1870	-	7471B	PPB	26 Apr 2021 11:53:56	U	SMPL
M2149-16 - 1	SB-40-9.5-10.0-DUP	881	0.1625	-	7471B	PPB	26 Apr 2021 11:56:12	U	SMPL
M2132-01LX5 - 1		13907	2.8286	-	7471B	PPB	26 Apr 2021 11:58:28	U	SMPL
M2132-01A - 1		87122	17.8144	-	7471B	PPB	26 Apr 2021 12:00:44	U	SMPL
M2149-01LX5 - 1		1199	0.2275	-	7471B	PPB	26 Apr 2021 12:03:00	U	SMPL
M2149-01A - 1		27772	5.6665	-	7471B	PPB	26 Apr 2021 12:05:20	U	SMPL
CCV94 - 1	CCV94	25398	5.1806	-	7471B	PPB	26 Apr 2021 12:07:36	U	SMPL
CCB94 - 1	CCB94	11	-0.0156	-	7471B	PPB	26 Apr 2021 12:09:53	U	SMPL
M2132-01DLX5 - 1		13577	2.7611	-	7471B	PPB	26 Apr 2021 12:12:09	U	SMPL
M2132-01DUPDLX5 - 1		13111	2.6657	-	7471B	PPB	26 Apr 2021 12:14:25	U	SMPL
M2132-01MSDLX5 - 1		16456	3.3504	-	7471B	PPB	26 Apr 2021 12:16:41	U	SMPL
M2132-01MSDDLX5 - 1		16802	3.4212	-	7471B	PPB	26 Apr 2021 12:18:56	U	SMPL
M2132-01LDLX25 - 1		2772	0.5495	-	7471B	PPB	26 Apr 2021 12:21:12	U	SMPL
M2132-01ADLX5 - 1		17530	3.5702	-	7471B	PPB	26 Apr 2021 12:23:29	U	SMPL
CCV95 - 1	CCV95	23718	4.8368	-	7471B	PPB	26 Apr 2021 12:25:45	U	SMPL
CCB95 - 1	CCB95	-5	-0.0189	-	7471B	PPB	26 Apr 2021 12:28:02	U	SMPL