



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114147

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/23/2021	17:36
2	CAL2	1	Water	NA	NA	20.3	7.00	04/23/2021	17:37
3	CAL3	1	Water	NA	NA	20.4	10.03	04/23/2021	17:38
4	ICV	1	Water	NA	NA	20.4	7.00	04/23/2021	17:39
5	CCV1	1	Water	NA	NA	20.4	2.01	04/23/2021	17:40
6	M2155-01	1	Solid	20.02	20	20.2	7.45	04/23/2021	17:42
7	M2159-01	1	Solid	20.03	20	20.4	10.05	04/23/2021	17:43
8	M2170-07	1	Solid	20.03	20	20.2	8.13	04/23/2021	17:51
9	M2170-07DUP	1	Solid	20.04	20	20.3	8.14	04/23/2021	17:52
10	M2170-09	1	Solid	20.03	20	20.3	9.76	04/23/2021	18:00
11	M2170-11	1	Solid	20.04	20	20.1	8.82	04/23/2021	18:01
12	M2170-13	1	Solid	20.02	20	20.3	8.69	04/23/2021	18:02
13	M2170-15	1	Solid	20.04	20	20.4	8.73	04/23/2021	18:05
14	M2170-17	1	Solid	20.01	20	20.6	9.84	04/23/2021	18:10
15	M2176-04	1	Solid	20.02	20	20.4	9.92	04/23/2021	18:13
16	CCV2	1	Water	NA	NA	20.4	12.02	04/23/2021	18:15
17	M2176-14	1	Solid	20.01	20	20.2	10.20	04/23/2021	18:19
18	M2176-18	1	Solid	20.03	20	20.3	10.17	04/23/2021	18:22
19	CCV3	1	Water	NA	NA	20.3	2.01	04/23/2021	18:24

WORKLIST(Hardcopy Internal Chain)

73 14147

WorkList Name : PH M2170

WorkList ID : 148333

Department : Wet-Chemistry

Date : 04-23-2021 14:20:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/23/2021	Solid	M2155-01	pH	Cool 4 deg C	PSEG03	O31	25123	04/23/2021	9045D
04/23/2021	Solid	M2159-01	pH	Cool 4 deg C	PSEG05	O32	BU-03-042321	04/23/2021	9045D
04/23/2021	Solid	M2170-07	pH	Cool 4 deg C	JPCL01	O31	DS-1	04/23/2021	9045D
04/23/2021	Solid	M2170-09	pH	Cool 4 deg C	JPCL01	O31	DN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-11	pH	Cool 4 deg C	JPCL01	O31	CS-1	04/23/2021	9045D
04/23/2021	Solid	M2170-13	pH	Cool 4 deg C	JPCL01	O31	CN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-15	pH	Cool 4 deg C	JPCL01	O31	BN-1	04/23/2021	9045D
04/23/2021	Solid	M2170-17	pH	Cool 4 deg C	JPCL01	O31	AN-1	04/23/2021	9045D
04/23/2021	Solid	M2176-04	pH	Cool 4 deg C	VERI01	P13	STOCKPILE-COMP-1	04/23/2021	9045D
04/23/2021	Solid	M2176-14	pH	Cool 4 deg C	VERI01	P13	STOCKPILE-COMP-3	04/23/2021	9045D
04/23/2021	Solid	M2176-18	pH	Cool 4 deg C	VERI01	P13	STOCKPILE-COMP-4	04/23/2021	9045D

Date/Time 04/23/21 14:30
 Raw Sample Received by: JO (wet chem)
 Raw Sample Relinquished by: J-L (Sun)

Date/Time 04/23/21 17:30
 Raw Sample Received by: J-L (Sun)
 Raw Sample Relinquished by: JO (wet chem)