



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114157

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/26/2021	13:05
2	M2154-02		1	15	100.00	>212.0	>212.0	04/26/2021	13:35
3	M2154-02DUP		1	15	100.00	>212.0	>212.0	04/26/2021	14:10
4	M2156-01		1	14	100.00	>212.0	>212.0	04/26/2021	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114157

WorkList Name : fp42621 WorkList ID : 148359 Department : Wet-Chemistry Date : 04-26-2021 09:28:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/29/2021	Water	M2154-02	C	Cool 4 deg C	PSEG03	O23	33096	04/23/2021	1010A
04/30/2021	Water	M2156-01	C	Cool 4 deg C	PSEG03	O31	33066	04/23/2021	1010A

Date/Time 04/26/21 12:25
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: RJ (Wetchem)

Date/Time 04/26/21 14:50
 Raw Sample Received by: RJ (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)