

LB114159


7470A

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

35583



A= 0.0000e+000

B= 2.7780e-004

C= -2.4983e-002

Rho= 0.9996863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	-0.012	-0.012	47	0.000	47					
0.2	0.200	0.186	-0.014	759	0.0 %	759					
2.5	2.500	2.443	-0.057	8884	0.0 %	8884					
5.0	5.000	5.072	0.072	18346	0.0 %	18346					
7.5	7.500	7.651	0.151	27633	0.0 %	27633					
10.0	10.000	9.860	-0.140	35583	0.0 %	35583					

LB114159 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Apr 2021 14:48:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	26 Apr 2021 14:56:15	S	Std
0.2 - 1	51.2	759	-	0.2000	7470A	PPB	26 Apr 2021 14:58:32	S	Std
2.5 - 1	52.5	8884	-	2.5000	7470A	PPB	26 Apr 2021 15:00:49	S	Std
5.0 - 1	55	18346	-	5.0000	7470A	PPB	26 Apr 2021 15:03:06	S	Std
7.5 - 1	57.5	27633	-	7.5000	7470A	PPB	26 Apr 2021 15:05:22	S	Std
10.0 - 1	510	35583	-	10.0000	7470A	PPB	26 Apr 2021 15:10:22	S	Std
ICV04 - 1	ICV04	13743	3.7928	-7470A	PPB	26 Apr 2021 15:13:10	U	SMPL	
ICB04 - 1	ICB04	0	-0.0250	-7470A	PPB	26 Apr 2021 15:17:56	U	SMPL	
CCV96 - 1	CCV96	18572	5.1343	-7470A	PPB	26 Apr 2021 15:20:10	U	SMPL	
CCB96 - 1	CCB96	5	-0.0236	-7470A	PPB	26 Apr 2021 15:24:56	U	SMPL	
CRA - 1	CRA	726	0.1767	-7470A	PPB	26 Apr 2021 15:27:11	U	SMPL	
HighStd - 1	HighStd	35199	9.7533	-7470A	PPB	26 Apr 2021 15:29:26	U	SMPL	
ChkStd - 1	ChkStd	25770	7.1339	-7470A	PPB	26 Apr 2021 15:31:41	U	SMPL	
PB135976BL - 1	PBW	99	0.0025	-7470A	PPB	26 Apr 2021 15:36:33	U	SMPL	
PB135976BS - 1	LCSW	15193	4.1956	-7470A	PPB	26 Apr 2021 15:38:48	U	SMPL	
M2132-02 - 1	TP-1-COMP	-81	-0.0475	-7470A	PPB	26 Apr 2021 15:41:05	U	SMPL	
M2132-02DUP - 1	TP-1-COMPDUP	33	-0.0158	-7470A	PPB	26 Apr 2021 15:43:21	U	SMPL	
M2132-02MS - 1	TP-1-COMPMS	15308	4.2276	-7470A	PPB	26 Apr 2021 15:45:38	U	SMPL	
M2132-02MSD - 1	TP-1-COMPMSD	15788	4.3609	-7470A	PPB	26 Apr 2021 15:47:55	U	SMPL	
M2132-07 - 1	TP-2-COMP	-139	-0.0636	-7470A	PPB	26 Apr 2021 15:50:11	U	SMPL	
CCV97 - 1	CCV97	18175	5.0241	-7470A	PPB	26 Apr 2021 15:52:27	U	SMPL	
CCB97 - 1	CCB97	-142	-0.0644	-7470A	PPB	26 Apr 2021 15:54:42	U	SMPL	
M2149-02 - 1	SB-36-COMP	15	-0.0208	-7470A	PPB	26 Apr 2021 15:56:58	U	SMPL	
M2149-05 - 1	SB-37-COMP	41	-0.0136	-7470A	PPB	26 Apr 2021 15:59:13	U	SMPL	
M2149-08 - 1	SB-38-COMP	63	-0.0075	-7470A	PPB	26 Apr 2021 16:01:29	U	SMPL	
M2149-11 - 1	SB-39-COMP	26	-0.0178	-7470A	PPB	26 Apr 2021 16:03:45	U	SMPL	
M2149-14 - 1	SB-40-COMP	60	-0.0083	-7470A	PPB	26 Apr 2021 16:06:01	U	SMPL	
M2152-02 - 1	TP-3-COMP	59	-0.0086	-7470A	PPB	26 Apr 2021 16:08:17	U	SMPL	
M2152-07 - 1	TP-4-COMP	62	-0.0078	-7470A	PPB	26 Apr 2021 16:10:33	U	SMPL	
PB135977BL - 1	PBW	47	-0.0119	-7470A	PPB	26 Apr 2021 16:12:50	U	SMPL	
PB135977BS - 1	LCSW	13402	3.6981	-7470A	PPB	26 Apr 2021 16:15:07	U	SMPL	
M2136-04 - 1	31075-76	377	0.0797	-7470A	PPB	26 Apr 2021 16:17:22	U	SMPL	
CCV98 - 1	CCV98	17588	4.8610	-7470A	PPB	26 Apr 2021 16:19:37	U	SMPL	
CCB98 - 1	CCB98	-139	-0.0636	-7470A	PPB	26 Apr 2021 16:21:53	U	SMPL	
M2136-04DUP - 1	31075-76DUP	398	0.0856	-7470A	PPB	26 Apr 2021 16:24:09	U	SMPL	
M2136-04MS - 1	31075-76MS	14132	3.9009	-7470A	PPB	26 Apr 2021 16:26:24	U	SMPL	
M2136-04MSD - 1	31075-76MSD	14335	3.9573	-7470A	PPB	26 Apr 2021 16:28:39	U	SMPL	
M2136-06 - 1	41765	-44	-0.0372	-7470A	PPB	26 Apr 2021 16:30:56	U	SMPL	
M2150-02 - 1	TWP-20	827	0.2048	-7470A	PPB	26 Apr 2021 16:33:13	U	SMPL	
M2150-03 - 1	TWP-20	-52	-0.0394	-7470A	PPB	26 Apr 2021 16:35:29	U	SMPL	
M2150-05 - 1	TWP-40	4475	1.2182	-7470A	PPB	26 Apr 2021 16:37:46	U	SMPL	
M2150-06 - 1	TWP-40	86	-0.0011	-7470A	PPB	26 Apr 2021 16:40:03	U	SMPL	
M2150-07 - 1	EQUIP-BLANK	35	-0.0153	-7470A	PPB	26 Apr 2021 16:42:19	U	SMPL	
M2150-08 - 1	EQUIP-BLANK	74	-0.0044	-7470A	PPB	26 Apr 2021 16:44:35	U	SMPL	
CCV99 - 1	CCV99	17549	4.8501	-7470A	PPB	26 Apr 2021 16:46:51	U	SMPL	
CCB99 - 1	CCB99	-63	-0.0425	-7470A	PPB	26 Apr 2021 16:49:06	U	SMPL	
PB135976TB - 1	PB135927TB	31	-0.0164	-7470A	PPB	26 Apr 2021 16:51:22	U	SMPL	
M2132-02LX5 - 1		31	-0.0164	-7470A	PPB	26 Apr 2021 16:53:37	U	SMPL	
M2132-02A - 1		14759	4.0751	-7470A	PPB	26 Apr 2021 16:55:52	U	SMPL	
M2136-04LX5 - 1		-12	-0.0283	-7470A	PPB	26 Apr 2021 16:58:08	U	SMPL	
M2136-04A - 1		14039	3.8751	-7470A	PPB	26 Apr 2021 17:00:24	U	SMPL	
CCV01 - 1	CCV01	17526	4.8438	-7470A	PPB	26 Apr 2021 17:02:40	U	SMPL	
CCB01 - 1	CCB01	-133	-0.0619	-7470A	PPB	26 Apr 2021 17:04:57	U	SMPL	