



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114163

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/26/2021	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/26/2021	16:11
3	CAL3	1	Water	NA	NA	20.2	10.03	04/26/2021	16:12
4	ICV	1	Water	NA	NA	20.3	7.01	04/26/2021	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	04/26/2021	16:18
6	M2187-05	1	Solid	20.03	20	20.5	8.70	04/26/2021	16:20
7	M2187-05DUP	1	Solid	20.04	20	20.6	8.71	04/26/2021	16:21
8	M2187-06	1	Solid	20.05	20	20.5	10.32	04/26/2021	16:23
9	CCV2	1	Water	NA	NA	20.4	12.02	04/26/2021	16:25

WORKLIST(Hardcopy Internal Chain)

20114163

WorkList Name : corrosivity m2187 WorkList ID : 148370 Department : Wet-Chemistry Date : 04-26-2021 15:22:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2021	Solid	M2187-05	Corrosivity	Cool 4 deg C	SKAN04	P21	TEST-PIT-I	04/26/2021	9045D
04/26/2021	Solid	M2187-06	Corrosivity	Cool 4 deg C	SKAN04	P21	TEST-PIT-I	04/26/2021	9045D

Date/Time 04/26/21 15:25
 Raw Sample Received by: J.R. (G.A. Ghem)
 Raw Sample Relinquished by: J.R. (G.A. Ghem)

Date/Time 04/26/21 16:30
 Raw Sample Received by: J.R. (G.A. Ghem)
 Raw Sample Relinquished by: J.R. (G.A. Ghem)