

LB114171


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

34329



A= 0.0000e+000

B= 2.8961e-004

C= 3.5202e-002

Rho= 0.9998243

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%01)
0.0	0.000	0.046	0.046	39	0.000	39					
0.2	0.200	0.229	0.029	669	0.0 %	669					
2.5	2.500	2.356	-0.144	8012	0.0 %	8012					
5.0	5.000	5.054	0.054	17328	0.0 %	17328					
7.5	7.500	7.538	0.038	25907	0.0 %	25907					
10.0	10.000	9.977	-0.023	34329	0.0 %	34329					

9/01)

15

-6

1

1

0

LB114171 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Apr 2021 09:58:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	27 Apr 2021 10:56:04	S	Std
0.2 - 1	50.2	669	-	0.2000	7470A	PPB	27 Apr 2021 10:58:21	S	Std
2.5 - 1	52.5	8012	-	2.5000	7470A	PPB	27 Apr 2021 11:00:37	S	Std
5.0 - 1	55	17328	-	5.0000	7470A	PPB	27 Apr 2021 11:02:54	S	Std
7.5 - 1	57.5	25907	-	7.5000	7470A	PPB	27 Apr 2021 11:05:10	S	Std
10.0 - 1	510	34329	-	10.0000	7470A	PPB	27 Apr 2021 11:10:42	S	Std
ICV05 - 1	ICV05	13875	4.0536	-7470A	PPB	27 Apr 2021 11:18:26	U	SMPL	
ICB05 - 1	ICB05	-173	-0.0149	-7470A	PPB	27 Apr 2021 11:20:40	U	SMPL	
CCV02 - 1	CCV02	18193	5.3041	-7470A	PPB	27 Apr 2021 11:22:56	U	SMPL	
CCB02 - 1	CCB02	-186	-0.0187	-7470A	PPB	27 Apr 2021 11:25:10	U	SMPL	
CRA - 1	CRA	594	0.2072	-7470A	PPB	27 Apr 2021 11:27:26	U	SMPL	
HighStd - 1	HighStd	34385	9.9935	-7470A	PPB	27 Apr 2021 11:29:41	U	SMPL	
ChkStd - 1	ChkStd	23992	6.9835	-7470A	PPB	27 Apr 2021 11:31:56	U	SMPL	
PB135994BL - 1	PBW	-251	-0.0375	-7470A	PPB	27 Apr 2021 11:34:14	U	SMPL	
PB135994BS - 1	LCSW	12289	3.5942	-7470A	PPB	27 Apr 2021 11:36:32	U	SMPL	
M2170-08 - 1	DS-1	32	0.0445	-7470A	PPB	27 Apr 2021 11:38:48	U	SMPL	
M2170-08DUP - 1	DS-1DUP	207	0.0952	-7470A	PPB	27 Apr 2021 11:41:05	U	SMPL	
M2170-08MS - 1	DS-1MS	12981	3.7946	-7470A	PPB	27 Apr 2021 11:43:22	U	SMPL	
M2170-08MSD - 1	DS-1MSD	12602	3.6849	-7470A	PPB	27 Apr 2021 11:45:38	U	SMPL	
M2170-10 - 1	DN-1	-131	-0.0027	-7470A	PPB	27 Apr 2021 11:47:54	U	SMPL	
CCV03 - 1	CCV03	17558	5.1202	-7470A	PPB	27 Apr 2021 11:50:10	U	SMPL	
CCB03 - 1	CCB03	-166	-0.0129	-7470A	PPB	27 Apr 2021 11:52:25	U	SMPL	
M2170-12 - 1	CS-1	30	0.0439	-7470A	PPB	27 Apr 2021 11:54:41	U	SMPL	
M2170-14 - 1	CN-1	17	0.0401	-7470A	PPB	27 Apr 2021 11:56:56	U	SMPL	
M2170-16 - 1	BN-1	169	0.0841	-7470A	PPB	27 Apr 2021 11:59:12	U	SMPL	
M2170-18 - 1	AN-1	79	0.0581	-7470A	PPB	27 Apr 2021 12:01:27	U	SMPL	
M2176-05 - 1	STOCKPILE-COMP-1	27	0.0430	-7470A	PPB	27 Apr 2021 12:03:43	U	SMPL	
M2176-10 - 1	STOCKPILE-COMP-2	33	0.0448	-7470A	PPB	27 Apr 2021 12:06:00	U	SMPL	
M2176-15 - 1	STOCKPILE-COMP-3	31	0.0442	-7470A	PPB	27 Apr 2021 12:08:16	U	SMPL	
M2176-19 - 1	STOCKPILE-COMP-4	45	0.0482	-7470A	PPB	27 Apr 2021 12:10:32	U	SMPL	
CCV04 - 1	CCV04	17323	5.0521	-7470A	PPB	27 Apr 2021 12:12:49	U	SMPL	
CCB04 - 1	CCB04	-132	-0.0030	-7470A	PPB	27 Apr 2021 12:15:05	U	SMPL	
PB135994TB - 1	PB135973TB	10	0.0381	-7470A	PPB	27 Apr 2021 12:17:22	U	SMPL	
M2170-08LX5 - 1		48	0.0491	-7470A	PPB	27 Apr 2021 12:19:37	U	SMPL	
M2170-08A - 1		13518	3.9502	-7470A	PPB	27 Apr 2021 12:21:52	U	SMPL	
CCV05 - 1	CCV05	17410	5.0773	-7470A	PPB	27 Apr 2021 12:24:07	U	SMPL	
CCB05 - 1	CCB05	-167	-0.0132	-7470A	PPB	27 Apr 2021 12:26:23	U	SMPL	