



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114172

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	04/27/2021	09:35
2	M2154-01		1	4	50.00	122.0	121.7	04/27/2021	10:10
3	M2154-01DUP		1	4	48.00	118.4	118.1	04/27/2021	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114172

WorkList Name : fp42621-1

WorkList ID : 148360

Department : Wet-Chemistry

Date : 04-26-2021 09:29:26

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/29/2021	Solid	M2154-01	✓	Flash Point	Cool 4 deg C	PSEG03	O23	32722	04/23/2021	1010A

Date/Time 04/27/21 09:05
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: RJ (EAL-100)

Date/Time 04/27/21 10:40
 Raw Sample Received by: RJ (EAL-100)
 Raw Sample Relinquished by: KS (Wetchem)