


LB114177

7471B

INSTRUMENT ID: CV1



| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | -0.022 | -0.022 | 50    | 0.000      | 50    |       |       |       |       | -9 |
| 0.2    | 0.200  | 0.182  | -0.018 | 1022  | 0.0 %      | 1022  |       |       |       |       | -2 |
| 2.5    | 2.500  | 2.461  | -0.039 | 11855 | 0.0 %      | 11855 |       |       |       |       | 1  |
| 5.0    | 5.000  | 5.037  | 0.037  | 24101 | 0.0 %      | 24101 |       |       |       |       | 3  |
| 7.5    | 7.500  | 7.741  | 0.241  | 36956 | 0.0 %      | 36956 |       |       |       |       | -2 |
| 10.0   | 10.000 | 9.800  | -0.200 | 46745 | 0.0 %      | 46745 |       |       |       |       |    |

## LB114177 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Apr 2021 13:34:01

| Sample ID       | Extended ID      | $\mu$ Abs. | Conc.   | Std Conc | Method | Units                | Date                 | Type | Type |
|-----------------|------------------|------------|---------|----------|--------|----------------------|----------------------|------|------|
| 0.0 - 1         | 50               | 50         | -       | 0.0000   | 7471B  | PPB                  | 27 Apr 2021 13:36:47 | S    | Std  |
| 0.2 - 1         | 50.2             | 1022       | -       | 0.2000   | 7471B  | PPB                  | 27 Apr 2021 13:39:04 | S    | Std  |
| 2.5 - 1         | 52.5             | 11855      | -       | 2.5000   | 7471B  | PPB                  | 27 Apr 2021 13:41:20 | S    | Std  |
| 5.0 - 1         | 55               | 24101      | -       | 5.0000   | 7471B  | PPB                  | 27 Apr 2021 13:43:37 | S    | Std  |
| 7.5 - 1         | 57.5             | 36956      | -       | 7.5000   | 7471B  | PPB                  | 27 Apr 2021 13:45:53 | S    | Std  |
| 10.0 - 1        | 510              | 46745      | -       | 10.0000  | 7471B  | PPB                  | 27 Apr 2021 13:50:45 | S    | Std  |
| ICV06 - 1       | ICV06            | 19026      | 3.9696  | -7471B   | PPB    | 27 Apr 2021 13:53:26 | U                    | SMPL |      |
| ICB06 - 1       | ICB06            | -73        | -0.0480 | -7471B   | PPB    | 27 Apr 2021 13:55:40 | U                    | SMPL |      |
| CCV06 - 1       | CCV06            | 24148      | 5.0470  | -7471B   | PPB    | 27 Apr 2021 13:57:57 | U                    | SMPL |      |
| CCB06 - 1       | CCB06            | 67         | -0.0186 | -7471B   | PPB    | 27 Apr 2021 14:02:45 | U                    | SMPL |      |
| CRA - 1         | CRA              | 1030       | 0.1840  | -7471B   | PPB    | 27 Apr 2021 14:04:59 | U                    | SMPL |      |
| HighStd - 1     | HighStd          | 47574      | 9.9748  | -7471B   | PPB    | 27 Apr 2021 14:07:14 | U                    | SMPL |      |
| ChkStd - 1      | ChkStd           | 34375      | 7.1983  | -7471B   | PPB    | 27 Apr 2021 14:09:30 | U                    | SMPL |      |
| PB136007BL - 1  | PBS              | 40         | -0.0243 | -7471B   | PPB    | 27 Apr 2021 14:11:48 | U                    | SMPL |      |
| PB136007BS - 1  | LCSS             | 19368      | 4.0415  | -7471B   | PPB    | 27 Apr 2021 14:14:06 | U                    | SMPL |      |
| M2159-01 - 1    | BU-03-042321     | 745        | 0.1240  | -7471B   | PPB    | 27 Apr 2021 14:16:22 | U                    | SMPL |      |
| M2159-01DUP - 1 | BU-03-042321DUP  | 895        | 0.1556  | -7471B   | PPB    | 27 Apr 2021 14:21:09 | U                    | SMPL |      |
| M2159-01MS - 1  | BU-03-042321MS   | 19118      | 3.9889  | -7471B   | PPB    | 27 Apr 2021 14:26:12 | U                    | SMPL |      |
| M2159-01MSD - 1 | BU-03-042321MSD  | 18826      | 3.9275  | -7471B   | PPB    | 27 Apr 2021 14:28:29 | U                    | SMPL |      |
| M2160-01 - 1    | OK-03-042321     | 3643       | 0.7336  | -7471B   | PPB    | 27 Apr 2021 14:30:44 | U                    | SMPL |      |
| CCV07 - 1       | CCV07            | 23834      | 4.9809  | -7471B   | PPB    | 27 Apr 2021 14:32:59 | U                    | SMPL |      |
| CCB07 - 1       | CCB07            | 61         | -0.0199 | -7471B   | PPB    | 27 Apr 2021 14:35:14 | U                    | SMPL |      |
| M2176-04 - 1    | STOCKPILE-COMP-1 | 2357       | 0.4631  | -7471B   | PPB    | 27 Apr 2021 14:37:30 | U                    | SMPL |      |
| M2176-09 - 1    | STOCKPILE-COMP-2 | 6959       | 1.4312  | -7471B   | PPB    | 27 Apr 2021 14:39:45 | U                    | SMPL |      |
| M2176-14 - 1    | STOCKPILE-COMP-3 | 10457      | 2.1670  | -7471B   | PPB    | 27 Apr 2021 14:42:00 | U                    | SMPL |      |
| M2176-18 - 1    | STOCKPILE-COMP-4 | 7178       | 1.4773  | -7471B   | PPB    | 27 Apr 2021 14:44:16 | U                    | SMPL |      |
| M2182-01 - 1    | TP-13-12-COMP    | 817        | 0.1392  | -7471B   | PPB    | 27 Apr 2021 14:46:31 | U                    | SMPL |      |
| M2183-01 - 1    | TP-16A-15-COMP   | 552        | 0.0834  | -7471B   | PPB    | 27 Apr 2021 14:48:48 | U                    | SMPL |      |
| M2183-03 - 1    | TP-14-COMP       | 950        | 0.1672  | -7471B   | PPB    | 27 Apr 2021 14:51:03 | U                    | SMPL |      |
| M2187-01 - 1    | TEST-PIT-I       | 27789      | 5.8129  | -7471B   | PPB    | 27 Apr 2021 14:53:20 | U                    | SMPL |      |
| M2187-03 - 1    | TEST-PIT-II      | 7545       | 1.5545  | -7471B   | PPB    | 27 Apr 2021 14:55:37 | U                    | SMPL |      |
| CCV08 - 1       | CCV08            | 24287      | 5.0762  | -7471B   | PPB    | 27 Apr 2021 14:57:53 | U                    | SMPL |      |
| CCB08 - 1       | CCB08            | -75        | -0.0485 | -7471B   | PPB    | 27 Apr 2021 15:00:09 | U                    | SMPL |      |
| M2159-01LX5 - 1 |                  | 178        | 0.0048  | -7471B   | PPB    | 27 Apr 2021 15:02:25 | U                    | SMPL |      |
| M2159-01A - 1   |                  | 19497      | 4.0686  | -7471B   | PPB    | 27 Apr 2021 15:04:40 | U                    | SMPL |      |
| CCV09 - 1       | CCV09            | 24213      | 5.0607  | -7471B   | PPB    | 27 Apr 2021 15:06:56 | U                    | SMPL |      |
| CCB09 - 1       | CCB09            | 28         | -0.0268 | -7471B   | PPB    | 27 Apr 2021 15:09:11 | U                    | SMPL |      |