


LB114121

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	0.010	0.010	25	0.000	25					-
0.05	0.050	0.064	0.014	206	0.0 %	206	0				28
0.2	0.200	0.196	-0.004	646	0.0 %	646					- 2
2.5	2.500	2.512	0.012	8411	0.0 %	8411					0
5.0	5.000	4.925	-0.075	16496	0.0 %	16496					- 2
7.5	7.500	7.535	0.035	25245	0.0 %	25245					0
10.0	10.000	10.007	0.007	33531	0.0 %	33531					0

LB114181 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Apr 2021 15:33:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	245.1	PPB	27 Apr 2021 15:39:20	S	Std
0.05 - 1	50.05	206	-	0.0500	245.1	PPB	27 Apr 2021 15:41:37	S	Std
0.2 - 1	50.2	646	-	0.2000	245.1	PPB	27 Apr 2021 15:43:54	S	Std
2.5 - 1	52.5	8411	-	2.5000	245.1	PPB	27 Apr 2021 15:46:11	S	Std
5.0 - 1	55	16496	-	5.0000	245.1	PPB	27 Apr 2021 15:48:27	S	Std
7.5 - 1	57.5	25245	-	7.5000	245.1	PPB	27 Apr 2021 15:56:33	S	Std
10.0 - 1	510	33531	-	10.0000	245.1	PPB	27 Apr 2021 16:07:15	S	Std
ICV07 - 1	ICV07	13667	4.0807	-245.1		PPB	27 Apr 2021 16:13:37	U	SMPL
ICB07 - 1	ICB07	7	0.0049	-245.1		PPB	27 Apr 2021 16:18:23	U	SMPL
CCV10 - 1	CCV10	17119	5.1106	-245.1		PPB	27 Apr 2021 16:23:13	U	SMPL
CCB10 - 1	CCB10	-147	-0.0410	-245.1		PPB	27 Apr 2021 16:25:28	U	SMPL
CRA - 1	CRA	204	0.0637	-245.1		PPB	27 Apr 2021 16:27:44	U	SMPL
HighStd - 1	HighStd	33947	10.1316	-245.1		PPB	27 Apr 2021 16:34:09	U	SMPL
ChkStd - 1	ChkStd	23467	7.0047	-245.1		PPB	27 Apr 2021 16:39:01	U	SMPL
PB136025BL - 1	PBW	-220	-0.0628	-245.1		PPB	27 Apr 2021 16:41:17	U	SMPL
PB136025BS - 1	LCSW	13596	4.0595	-245.1		PPB	27 Apr 2021 16:46:05	U	SMPL
M2150-01 - 1	TWP-20	1798	0.5393	-245.1		PPB	27 Apr 2021 16:48:21	U	SMPL
M2150-04 - 1	TWP-40	4571	1.3667	-245.1		PPB	27 Apr 2021 16:50:38	U	SMPL
M2150-04DUP - 1	TWP-40DUP	4184	1.2512	-245.1		PPB	27 Apr 2021 16:52:55	U	SMPL
M2150-04MS - 1	TWP-40MS	17715	5.2884	-245.1		PPB	27 Apr 2021 16:55:11	U	SMPL
M2150-04MSD - 1	TWP-40MSD	17728	5.2923	-245.1		PPB	27 Apr 2021 16:57:27	U	SMPL
CCV11 - 1	CCV11	16758	5.0029	-245.1		PPB	27 Apr 2021 16:59:41	U	SMPL
CCB11 - 1	CCB11	-80	-0.0210	-245.1		PPB	27 Apr 2021 17:01:57	U	SMPL
M2184-03 - 1	NCQ-632	841	0.2538	-245.1		PPB	27 Apr 2021 17:04:12	U	SMPL
M2150-04LX5 - 1		832	0.2511	-245.1		PPB	27 Apr 2021 17:06:27	U	SMPL
M2150-04A - 1		18534	5.5328	-245.1		PPB	27 Apr 2021 17:08:42	U	SMPL
CCV12 - 1	CCV12	16590	4.9528	-245.1		PPB	27 Apr 2021 17:10:58	U	SMPL
CCB12 - 1	CCB12	-74	-0.0192	-245.1		PPB	27 Apr 2021 17:13:13	U	SMPL