



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114183

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	04/27/2021	16:41
2	CAL2	1	Water	NA	NA	20.3	7.00	04/27/2021	16:42
3	CAL3	1	Water	NA	NA	20.3	10.03	04/27/2021	16:43
4	ICV	1	Water	NA	NA	20.4	7.01	04/27/2021	16:45
5	CCV1	1	Water	NA	NA	20.3	2.01	04/27/2021	16:47
6	M2196-02	1	Solid	20.02	20	20.7	6.89	04/27/2021	16:48
7	M2196-06	1	Solid	20.02	20	20.6	6.93	04/27/2021	16:49
8	M2198-02	1	Solid	20.03	20	20.6	7.52	04/27/2021	16:51
9	M2199-04	1	Solid	20.02	20	20.5	8.17	04/27/2021	16:55
10	M2199-04DUP	1	Solid	20.03	20	20.6	8.18	04/27/2021	16:56
11	CCV2	1	Water	NA	NA	20.4	12.02	04/27/2021	17:00

WORKLIST(Hardcopy Internal Chain)

28 114183

WorkList Name : corrosivity m2196

WorkList ID : 148404

Department : Wet-Chemistry

Date : 04-27-2021 14:50:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/27/2021	Solid	M2196-02	Corrosivity	Cool 4 deg C	TULL02	P22	WALL-D-4075-A	04/27/2021	9045D
04/27/2021	Solid	M2196-06	Corrosivity	Cool 4 deg C	TULL02	P22	WALL-D-4075-B	04/27/2021	9045D
04/26/2021	Solid	M2198-02	Corrosivity	Cool 4 deg C	LIRO01	P31	SAMPLE-2-COMP	04/26/2021	9045D
04/26/2021	Solid	M2199-04	Corrosivity	Cool 4 deg C	LIRO01	P22	SAMPLE-3-COMP	04/26/2021	9045D

Date/Time 04/27/21 15:00
 Raw Sample Received by: JP (not 12pm)
 Raw Sample Relinquished by: JC (5pm)

Date/Time 04/27/21 17:00
 Raw Sample Received by: JC (5pm)
 Raw Sample Relinquished by: JP (not 12pm)