

LB114190

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : Am

Instrument ID : Konelab

4/28/2021 14:24

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.000	0.0	0.106	
ICB1	-0.007	0.0	0.005	
CCV1	1.002	0.0	0.106	
CCB1	-0.004	0.0	0.005	
PB135953BL	-0.005	0.0	0.005	
PB135953BS	1.000	0.0	0.106	
M2150-01	0.057	0.0	0.011	
M2150-01DUP	0.061	0.0	0.011	
M2150-01MS	1.002	0.0	0.106	
M2150-01MSD	1.000	0.0	0.105	
M2150-04	0.031	0.0	0.008	
CCV2	0.994	0.0	0.105	
CCB2	-0.008	0.0	0.005	

N	13
Mean	0.471
SD	0.5100
CV%	108.30

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 28 11:56:31 2021

Wed Apr 28 14:14:41 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	-0.0031	mg/l	4/28/2021 11:56:31	
0.05PPM	A	Phenolics-	P	0.052	mg/l	4/28/2021 11:56:32	
0.1PPM	A	Phenolics-	P	0.107	mg/l	4/28/2021 11:56:33	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	4/28/2021 11:56:34	
0.50PPM	A	Phenolics-	P	0.4989	mg/l	4/28/2021 11:56:35	
1.0PPM	A	Phenolics-	P	0.9912	mg/l	4/28/2021 11:56:36	
2.0PPM	A	Phenolics-	P	2.0043	mg/l	4/28/2021 11:56:37	
ICV1	S	Phenolics-	P	0.9995	mg/l	4/28/2021 13:12:43	
ICB1	S	Phenolics-	P	-0.0072	mg/l	4/28/2021 13:12:46	
CCV1	S	Phenolics-	P	1.0021	mg/l	4/28/2021 13:12:48	
CCB1	S	Phenolics-	P	-0.0044	mg/l	4/28/2021 13:12:49	
PB135953BL	S	Phenolics-	P	-0.0047	mg/l	4/28/2021 13:12:52	
PB135953BS	S	Phenolics-	P	1.0005	mg/l	4/28/2021 13:12:53	
M2150-01	S	Phenolics-	P	0.0568	mg/l	4/28/2021 13:22:45	
M2150-01DUP	S	Phenolics-	P	0.0609	mg/l	4/28/2021 13:22:48	
M2150-01MS	S	Phenolics-	P	1.0016	mg/l	4/28/2021 13:22:50	
M2150-01MSD	S	Phenolics-	P	0.9995	mg/l	4/28/2021 13:22:52	
M2150-04	S	Phenolics-	P	0.0308	mg/l	4/28/2021 13:22:54	
CCV2	S	Phenolics-	P	0.9944	mg/l	4/28/2021 13:22:56	
CCB2	S	Phenolics-	P	-0.0082	mg/l	4/28/2021 13:26:43	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

4/28/2021 12:49

Reviewed by : AM Instrument ID : Konelab


Test Phenolics-

Accepted 4/28/2021 11:57

Factor 9.979
Bias 0.005

Coeff. of det. 0.999949

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %RE
1	0.00PPM	0.005	-0.0031	0.0000	-
2	Phenol-2	0.011	0.0520	0.0500	4.0
3	Phenol-2	0.016	0.1070	0.1000	7.0
4	Phenol-2	0.030	0.2497	0.2500	-0.1
5	Phenol-2	0.055	0.4989	0.5000	-0.2
6	Phenol-2	0.105	0.9912	1.0000	-0.9
7	Phenol-2	0.206	2.0043	2.0000	0.2

04/28/2021
AM