



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114200

Slope : 99.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/28/2021	15:20
2	CAL2	1	Water	NA	NA	20.4	7.00	04/28/2021	15:21
3	CAL3	1	Water	NA	NA	20.4	10.04	04/28/2021	15:22
4	ICV	1	Water	NA	NA	20.5	6.99	04/28/2021	15:25
5	CCV1	1	Water	NA	NA	20.3	2.01	04/28/2021	15:26
6	M2202-03	1	Water	NA	NA	20.2	12.26	04/28/2021	15:28
7	M2202-03DUP	1	Water	NA	NA	20.3	12.27	04/28/2021	15:29
8	M2202-04	1	Water	NA	NA	20.3	12.23	04/28/2021	15:30
9	M2202-05	1	Water	NA	NA	20.3	12.27	04/28/2021	15:32
10	M2202-06	1	Water	NA	NA	20.4	12.27	04/28/2021	15:33
11	M2202-07	1	Water	NA	NA	20.4	12.22	04/28/2021	15:35
12	CCV2	1	Water	NA	NA	20.3	12.02	04/28/2021	15:36

114200

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2204 WorkList ID : 148435 Department : Wet-Chemistry Date : 04-28-2021 15:07:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Water	M2202-03 ✓	pH	Cool 4 deg C	PSEG03	O41	TOTE-A	04/28/2021	9040C
04/28/2021	Water	M2202-04 ✓	pH	Cool 4 deg C	PSEG03	O41	TOTE-B	04/28/2021	9040C
04/28/2021	Water	M2202-05 ✓	pH	Cool 4 deg C	PSEG03	O41	TOTE-C	04/28/2021	9040C
04/28/2021	Water	M2202-06 ✓	pH	Cool 4 deg C	PSEG03	O41	TOTE-D	04/28/2021	9040C
04/28/2021	Water	M2202-07 ✓	pH	Cool 4 deg C	PSEG03	O41	TOTE-E	04/28/2021	9040C

Date/Time 04/28/21 15:10
 Raw Sample Received by: JP Wet Chem
 Raw Sample Relinquished by: JDCSM

Date/Time 04/28/21 16:30
 Raw Sample Received by: J.D.CSM
 Raw Sample Relinquished by: JP Wet Chem