



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114202

Slope : 99.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/28/2021	15:20
2	CAL2	1	Water	NA	NA	20.4	7.00	04/28/2021	15:21
3	CAL3	1	Water	NA	NA	20.4	10.04	04/28/2021	15:22
4	ICV	1	Water	NA	NA	20.5	6.99	04/28/2021	15:25
5	CCV1	1	Water	NA	NA	20.3	2.01	04/28/2021	15:26
6	M2219-01	1	Solid	20.03	20	23.6	6.53	04/28/2021	16:22
7	M2219-01DUP	1	Solid	20.04	20	23.7	6.54	04/28/2021	16:23
8	M2223-01	1	Solid	20.02	20	24.2	7.20	04/28/2021	16:25
9	M2223-03	1	Solid	20.03	20	23.9	7.14	04/28/2021	16:26
10	CCV2	1	Water	NA	NA	20.2	12.02	04/28/2021	16:27

WORKLIST(Hardcopy Internal Chain)

AB 11/202

WorkList Name : PH M2223 S WorkList ID : 148439 Department : Wet-Chemistry Date : 04-28-2021 17:01:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Solid	M2219-01	pH	Cool 4 deg C	TULL02	P31	FK-PS-01-071	04/28/2021	9045D
04/27/2021	Solid	M2223-01	pH	Cool 4 deg C	METE01	P23	LOADING-AREA-MUD	04/27/2021	9045D
04/27/2021	Solid	M2223-03	pH	Cool 4 deg C	METE01	P23	LOADING-AREA-PAD	04/27/2021	9045D

Date/Time 04/28/21 17:10
 Raw Sample Received by: JP Cef-B-2022
 Raw Sample Relinquished by: Amy Wetchem

Date/Time 04/28/21 18:20
 Raw Sample Received by: Amy Wetchem
 Raw Sample Relinquished by: JP Cef-B-2022