



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114205

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	04/29/2021	11:25
2	CAL2	1	Water	NA	NA	20.6	7.00	04/29/2021	11:26
3	CAL3	1	Water	NA	NA	20.6	10.03	04/29/2021	11:27
4	ICV	1	Water	NA	NA	20.5	6.98	04/29/2021	11:28
5	CCV1	1	Water	NA	NA	20.4	2.01	04/29/2021	11:30
6	M2176-09	1	Solid	20.02	20	20.8	10.54	04/29/2021	11:32
7	M2176-09DUP	1	Solid	20.03	20	20.7	10.55	04/29/2021	11:33
8	CCV2	1	Water	NA	NA	20.5	12.02	04/29/2021	11:35

VB 114205

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2176 WorkList ID : 148464 Department : Wet-Chemistry Date : 04-29-2021 11:20:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/23/2021	Solid	M2176-09	pH	Cool 4 deg C	VERI01	P13	STOCKPILE-COMP-2	04/23/2021	9045D

Date/Time 04/29/21 11:22
 Raw Sample Received by: JP (wet Chem)
 Raw Sample Relinquished by: JP (SM)

Date/Time 04/29/21 12:00
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (wet Chem)