



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114207

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	04/29/2021	11:30
2	CAL2	1	Water	NA	NA	21.1	7.01	04/29/2021	11:31
3	CAL3	1	Water	NA	NA	21.1	10.03	04/29/2021	11:32
4	ICV	1	Water	NA	NA	21.2	7.00	04/29/2021	11:35
5	CCV1	1	Water	NA	NA	21.1	2.01	04/29/2021	11:36
6	M2228-01	1	Water	NA	NA	20.7	5.07	04/29/2021	11:40
7	M2228-01DUP	1	Water	NA	NA	20.8	5.08	04/29/2021	11:41
8	M2228-02	1	Water	NA	NA	20.6	4.89	04/29/2021	11:43
9	M2228-03	1	Water	NA	NA	20.4	5.17	04/29/2021	11:45
10	CCV2	1	Water	NA	NA	21.3	12.02	04/29/2021	11:46

WB 114207

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2228 WATER

WorkList ID : 148465

Department : Wet-Chemistry

Date : 04-29-2021 11:21:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Water	M2228-01	pH	Cool 4 deg C	STER09	P31	CELL-A -POOL-WATER(I	04/28/2021	9040C
04/28/2021	Water	M2228-02	pH	Cool 4 deg C	STER09	P31	CELL-B-C-POOL-WATER	04/28/2021	9040C
04/28/2021	Water	M2228-03	pH	Cool 4 deg C	STER09	P31	CELL-D-POOL-WATER-(I	04/28/2021	9040C

Date/Time 04/29/21 11:25
Raw Sample Received by: JP (wet Chem)
Raw Sample Relinquished by: CR (SM)

Date/Time 04/29/21 12:10
Raw Sample Received by: OP (SM)
Raw Sample Relinquished by: JP (wet Chem)