


LB114208

7470A

INSTRUMENT ID: CV1

 μ Abs.:

38476



Linear

A= 0.0000e+000

B= 2.6037e-004

C= -2.5122e-002

Rho= 0.9999981

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	113	0.000	113					
0.2	0.200	0.194	-0.006	842	0.0 %	842					
2.5	2.500	2.501	0.001	9703	0.0 %	9703					
5.0	5.000	4.995	-0.005	19280	0.0 %	19280					
7.5	7.500	7.513	0.013	28952	0.0 %	28952					
10.0	10.000	9.993	-0.007	38476	0.0 %	38476					

0/00
-3
0
0
0
0

LB114208 INSTRUMENT ID :: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Apr 2021 11:57:32

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50.2	113	-	0.00007470A	PPB	29 Apr 2021 12:04:26	S	Std
0.2 - 1	50.2	842	-	0.20007470A	PPB	29 Apr 2021 12:06:43	S	Std
2.5 - 1	52.5	9703	-	2.50007470A	PPB	29 Apr 2021 12:09:00	S	Std
5.0 - 1	53.5	19280	-	5.00007470A	PPB	29 Apr 2021 12:11:16	S	Std
7.5 - 1	53.5	28952	-	7.50007470A	PPB	29 Apr 2021 12:13:33	S	Std
10.0 - 1	50	38476	-	10.00007470A	PPB	29 Apr 2021 12:15:50	S	Std
ICV09 - 1	ICV09	15698	4.0621	-7470A	PPB	29 Apr 2021 12:20:59	U	SMPL
ICB09 - 1	ICB09	59	-0.0098	-7470A	PPB	29 Apr 2021 12:23:14	U	SMPL
CCV16 - 1	CCV16	19883	5.1517	-7470A	PPB	29 Apr 2021 12:25:29	U	SMPL
CCB16 - 1	CCB16	27	-0.0181	-7470A	PPB	29 Apr 2021 12:27:44	U	SMPL
CRA - 1	CRA	816	0.1873	-7470A	PPB	29 Apr 2021 12:30:00	U	SMPL
HighStd - 1	HighStd	38512	10.0021	-7470A	PPB	29 Apr 2021 12:32:14	U	SMPL
ChkStd - 1	ChkStd	27496	7.1339	-7470A	PPB	29 Apr 2021 12:34:29	U	SMPL
PB136045BL - 1	PBW	-12	-0.0282	-7470A	PPB	29 Apr 2021 12:36:47	U	SMPL
PB136045BS - 1	LCSW	14115	3.6499	-7470A	PPB	29 Apr 2021 12:39:04	U	SMPL
M2175-01 - 1	BACKFILL-ASHCROFT-REAR-03	589	0.1282	-7470A	PPB	29 Apr 2021 12:46:13	U	SMPL
M2175-01DUP - 1	BACKFILL-ASHCROFT-REAR-03DUP	482	0.1004	-7470A	PPB	29 Apr 2021 12:48:29	U	SMPL
M2175-01MS - 1	BACKFILL-ASHCROFT-REAR-03MS	14156	3.6606	-7470A	PPB	29 Apr 2021 12:50:46	U	SMPL
M2175-01MSD - 1	BACKFILL-ASHCROFT-REAR-03MSD	14903	3.8551	-7470A	PPB	29 Apr 2021 12:53:02	U	SMPL
M2175-02 - 1	BACKFILL-ASHCROFT-REAR-04	358	0.0681	-7470A	PPB	29 Apr 2021 12:55:17	U	SMPL
CCV17 - 1	CCV17	19707	5.1059	-7470A	PPB	29 Apr 2021 12:57:32	U	SMPL
CCB17 - 1	CCB17	-53	-0.0389	-7470A	PPB	29 Apr 2021 12:59:48	U	SMPL
M2175-03 - 1	BACKFILL-ASHCROFT-REAR-05	543	0.1163	-7470A	PPB	29 Apr 2021 13:02:03	U	SMPL
PB136046BL - 1	PBW	-20	-0.0303	-7470A	PPB	29 Apr 2021 13:04:18	U	SMPL
PB136046BS - 1	LCSW	15241	3.9431	-7470A	PPB	29 Apr 2021 13:06:34	U	SMPL
M2196-02 - 1	WALL-D-4075-A	241	0.0376	-7470A	PPB	29 Apr 2021 13:08:49	U	SMPL
M2196-06 - 1	WALL-D-4075-B	268	0.0447	-7470A	PPB	29 Apr 2021 13:11:05	U	SMPL
M2198-02 - 1	SAMPLE-2-COMP	283	0.0486	-7470A	PPB	29 Apr 2021 13:13:21	U	SMPL
M2199-04 - 1	SAMPLE-3-COMP	275	0.0465	-7470A	PPB	29 Apr 2021 13:15:37	U	SMPL
M2199-04DUP - 1	SAMPLE-3-COMPDUP	272	0.0457	-7470A	PPB	29 Apr 2021 13:17:54	U	SMPL
M2199-04MS - 1	SAMPLE-3-COMPMS	13281	3.4328	-7470A	PPB	29 Apr 2021 13:20:11	U	SMPL
M2199-04MSD - 1	SAMPLE-3-COMPMSD	14984	3.8762	-7470A	PPB	29 Apr 2021 13:22:26	U	SMPL
CCV18 - 1	CCV18	19546	5.0640	-7470A	PPB	29 Apr 2021 13:24:42	U	SMPL
CCB18 - 1	CCB18	-55	-0.0394	-7470A	PPB	29 Apr 2021 13:26:57	U	SMPL
PB136045TB - 1	PB135990TB	240	0.0374	-7470A	PPB	29 Apr 2021 13:29:12	U	SMPL
PB136046TB - 1	PB136018TB	313	0.0564	-7470A	PPB	29 Apr 2021 13:31:27	U	SMPL
M2175-01LX5 - 1		279	0.0475	-7470A	PPB	29 Apr 2021 13:33:43	U	SMPL
M2175-01A - 1		15583	4.0322	-7470A	PPB	29 Apr 2021 13:35:58	U	SMPL
M2199-04LX5 - 1		-49	-0.0379	-7470A	PPB	29 Apr 2021 13:38:14	U	SMPL
M2199-04A - 1		15691	4.0603	-7470A	PPB	29 Apr 2021 13:40:30	U	SMPL
CCV19 - 1	CCV19	19846	5.1421	-7470A	PPB	29 Apr 2021 13:42:47	U	SMPL
CCB19 - 1	CCB19	-67	-0.0426	-7470A	PPB	29 Apr 2021 13:45:03	U	SMPL