



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114215

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/29/2021	14:10
2	CAL2	1	Water	NA	NA	20.7	7.00	04/29/2021	14:11
3	CAL3	1	Water	NA	NA	20.6	10.03	04/29/2021	14:12
4	ICV	1	Water	NA	NA	20.6	6.99	04/29/2021	14:15
5	CCV1	1	Water	NA	NA	20.6	2.02	04/29/2021	14:17
6	M2231-01	1	Water	NA	NA	20.6	5.02	04/29/2021	14:19
7	M2231-01DUP	1	Water	NA	NA	20.7	5.03	04/29/2021	14:20
8	CCV2	1	Water	NA	NA	20.3	12.02	04/29/2021	14:21

18114215

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2231 WorkList ID : 148471 Department : Wet-Chemistry Date : 04-29-2021 13:56:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/09/2021	Water	M2231-01	✓ pH	Cool 4 deg C	STER14	P51	MINI-CELL	04/09/2021	9040C

Date/Time 04/29/21 14:00
Raw Sample Received by: JP ~~Wet Chem~~
Raw Sample Relinquished by: CP (SM)

Date/Time 04/29/21 15:00
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: JP ~~Wet Chem~~