

LB114216


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53369



A= 0.0000e+000

B= 1.8971e-004

C= 1.4864e-002

Rho= 0.9995780

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.030	0.030	80	0.000	80					
0.2	0.200	0.230	0.030	1136	0.0 %	1136					15
2.5	2.500	2.475	-0.025	12969	0.0 %	12969					-1
5.0	5.000	5.038	0.038	26477	0.0 %	26477					1
7.5	7.500	7.287	-0.213	38330	0.0 %	38330					-3
10.0	10.000	10.140	0.140	53369	0.0 %	53369					1

LB114216 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Apr 2021 15:18:34					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	80	-	0.0000	7471B	PPB	29 Apr 2021 15:22:10	S	Std
0.2 - 1	SD.2	1136	-	0.2000	7471B	PPB	29 Apr 2021 15:24:26	S	Std
2.5 - 1	SD.5	12969	-	2.5000	7471B	PPB	29 Apr 2021 15:26:44	S	Std
5.0 - 1	SD.5	26477	-	5.0000	7471B	PPB	29 Apr 2021 15:29:01	S	Std
7.5 - 1	SD.5	38330	-	7.5000	7471B	PPB	29 Apr 2021 15:33:48	S	Std
10.0 - 1	SD.10	53369	-	10.0000	7471B	PPB	29 Apr 2021 15:41:15	S	Std
ICV10 - 1	ICV10	20350	3.8756	-	7471B	PPB	29 Apr 2021 15:46:26	U	SMPL
ICB10 - 1	ICB10	24	0.0194	-	7471B	PPB	29 Apr 2021 15:48:40	U	SMPL
CCV20 - 1	CCV20	25917	4.9317	-	7471B	PPB	29 Apr 2021 15:50:56	U	SMPL
CCB20 - 1	CCB20	-6	0.0137	-	7471B	PPB	29 Apr 2021 15:53:11	U	SMPL
CRA - 1	CRA	1088	0.2213	-	7471B	PPB	29 Apr 2021 15:55:27	U	SMPL
HighStd - 1	HighStd	53654	10.1938	-	7471B	PPB	29 Apr 2021 15:57:41	U	SMPL
ChkStd - 1	ChkStd	35641	6.7765	-	7471B	PPB	29 Apr 2021 15:59:57	U	SMPL
PB136067BL - 1	PBS	-48	0.0058	-	7471B	PPB	29 Apr 2021 16:02:14	U	SMPL
PB136067BS - 1	LCSS	20682	3.9385	-	7471B	PPB	29 Apr 2021 16:04:31	U	SMPL
M2202-01 - 1	290	603	0.1293	-	7471B	PPB	29 Apr 2021 16:06:47	U	SMPL
M2207-01 - 1	TP-6A-5-COMP	3751	0.7265	-	7471B	PPB	29 Apr 2021 16:09:04	U	SMPL
M2207-03 - 1	TP-4A-3-COMP	1742	0.3453	-	7471B	PPB	29 Apr 2021 16:11:21	U	SMPL
M2217-15 - 1	HTCRV2-GRAB-CF02-01	471	0.1042	-	7471B	PPB	29 Apr 2021 16:13:38	U	SMPL
M2217-15DUP - 1	HTCRV2-GRAB-CF02-01DUP	551	0.1194	-	7471B	PPB	29 Apr 2021 16:18:25	U	SMPL
CCV21 - 1	CCV21	25992	4.9459	-	7471B	PPB	29 Apr 2021 16:20:39	U	SMPL
CCB21 - 1	CCB21	-93	-0.0028	-	7471B	PPB	29 Apr 2021 16:22:54	U	SMPL
M2217-16 - 1	M2217-15MS	21737	4.1387	-	7471B	PPB	29 Apr 2021 16:25:09	U	SMPL
M2217-17 - 1	M2217-15MSD	19340	3.6839	-	7471B	PPB	29 Apr 2021 16:27:25	U	SMPL
M2217-18 - 1	HTCRV2-GRAB-CF02-02	447	0.0997	-	7471B	PPB	29 Apr 2021 16:29:40	U	SMPL
M2217-19 - 1	HTCRV2-GRAB-TS01-01	4432	0.8557	-	7471B	PPB	29 Apr 2021 16:31:56	U	SMPL
M2217-20 - 1	M2217-19MS	22704	4.3221	-	7471B	PPB	29 Apr 2021 16:34:12	U	SMPL
M2217-21 - 1	M2217-19MSD	24573	4.6767	-	7471B	PPB	29 Apr 2021 16:36:28	U	SMPL
M2217-22 - 1	HTCRV2-GRAB-TS01-02	4233	0.8179	-	7471B	PPB	29 Apr 2021 16:38:45	U	SMPL
M2218-01 - 1	FK-BPM-01-C6-D6-C6A-D6A	1136	0.2304	-	7471B	PPB	29 Apr 2021 16:41:02	U	SMPL
M2219-01 - 1	FK-PS-01-071	1628	0.3237	-	7471B	PPB	29 Apr 2021 16:43:19	U	SMPL
M2222-01 - 1	TP-5-COMP	7734	1.4821	-	7471B	PPB	29 Apr 2021 16:45:34	U	SMPL
CCV22 - 1	CCV22	26301	5.0045	-	7471B	PPB	29 Apr 2021 16:47:50	U	SMPL
CCB22 - 1	CCB22	-69	0.0018	-	7471B	PPB	29 Apr 2021 16:50:05	U	SMPL
M2222-06 - 1	TP-5A-COMP	8990	1.7204	-	7471B	PPB	29 Apr 2021 16:52:21	U	SMPL
M2217-15LX5 - 1		48	0.0240	-	7471B	PPB	29 Apr 2021 16:54:36	U	SMPL
M2217-15A - 1		26025	4.9522	-	7471B	PPB	29 Apr 2021 16:59:24	U	SMPL
CCV23 - 1	CCV23	25665	4.8839	-	7471B	PPB	29 Apr 2021 17:01:40	U	SMPL
CCB23 - 1	CCB23	-79	-0.0001	-	7471B	PPB	29 Apr 2021 17:03:56	U	SMPL