



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114222

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/30/2021	10:32
2	CAL2	1	Water	NA	NA	20.2	7.00	04/30/2021	10:33
3	CAL3	1	Water	NA	NA	20.3	10.03	04/30/2021	10:35
4	ICV	1	Water	NA	NA	20.4	7.00	04/30/2021	10:36
5	CCV1	1	Water	NA	NA	20.3	2.01	04/30/2021	10:39
6	M2213-04	1	Water	NA	NA	20.7	5.92	04/30/2021	10:40
7	M2213-04DUP	1	Water	NA	NA	20.8	5.93	04/30/2021	10:42
8	CCV2	1	Water	NA	NA	20.4	12.02	04/30/2021	10:43

WORKLIST(Hardcopy Internal Chain)

XB 114222

WorkList Name : PH M2213 WorkList ID : 148499 Department : Wet-Chemistry Date : 04-30-2021 10:28:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/28/2021	Water	M2213-04	pH	Cool 4 deg C	STER12	P31	CELL-POOL-WATER-2N	04/28/2021	9040C

Date/Time 04/30/21 10:30
 Raw Sample Received by: JD (404-2em)
 Raw Sample Relinquished by: (AKSM)

Date/Time 04/30/21 11:20
 Raw Sample Received by: (AKSM)
 Raw Sample Relinquished by: JD (404-2em)